



Baltic Fire Laboratory
Ul. Banińska 8
80-209 Tuchom
Poland
email: info@bafilab.com

TEST REPORT

No. TRBFL2525-1

Customer: Plumis Ltd, Unit 4, Phoenix Trading Estate Bilton Road Perivale London, United Kingdom, UB6 7DZ

Producer: Plumis Ltd, Unit 4, Phoenix Trading Estate Bilton Road Perivale London, United Kingdom, UB6 7DZ

Test Method: BS EN 14972-17:2025 „Fixed firefighting systems – Water mist systems – Part 17: Test protocol for residential occupancies for automatic nozzle systems”

Date of testing & place: 25– 27 November 2025 at Baltic Fire Laboratory, Tuchom, Poland



Test report prepared by: Magdalena Siedzielnik, Researcher (BFL)

Magdalena Siedzielnik

Test report verified & authorized by: Bogdan Racięga, Fire Laboratory Director (BFL)

Bogdan Racięga

Report issue date and place:

Tuchom, March 18, 2026



ABSTRACT

Plumis initiated a test programme to assess the performance of its residential water mist system using the fire scenarios defined in BS EN 14972-17:2025.

As electronically actuated nozzles fall outside the scope of BS EN 14972-17:2025, formal compliance with Part 17 cannot be claimed for this system configuration.

To provide a relevant performance assessment, the Part 17 fire test scenarios were applied with a defined deviation whereby activation of the water mist nozzles was initiated by a fire detector.

The system utilised sidewall nozzles installed 1.2 m above the floor, spaced 4 m apart and positioned 2 m from the wall within an enclosure measuring 8 × 4 × 2.5 m (L × W × H). The objective of the testing was to verify suppression performance under the defined fire scenarios.

The water mist nozzles were received at Baltic Fire Laboratory, Tuchom, Poland on 17 November 2025. Non-accredited R&D testing was conducted on 25– 27 November 2025 at Baltic Fire Laboratory.

The purpose of the test was to evaluate the firefighting performance of an installed water mist system:

- SH11

Six tests were performed accordance with BS EN 14972-17:2025. Three tests were carried out in a 8.0 x 4.0 [m] test enclosure with a ceiling height of 2.5 [m]. One test was performed in enclosed test enclosure with a ceiling height of 3.5 [m]. One test were performed in open space (corner test) with a ceiling height of 2.5 [m]. In all fire tests, the nozzles were installed on the wall at a height of 1.2 [m] above the floor.

The tested water mist system met the criteria of BS EN 14972-17:2025. The third nozzle, located outside the room, did not operate in any of the tests. The temperature values presented in Table 1 below were not exceeded.

Table 1. Fire test maximum temperatures.

Thermocouple location	Maximum allowable temperature [°C]
75 [mm] below the underside of the ceiling	320
1.6 [m] above the floor	95
1.6 [m] above the floor	55*
*For not more than any 120s interval.	

Specification of the nozzle supplied by Plumis is presented in Table 2.

Table 2. Water mist nozzle characteristic SH11.

WATER MIST NOZZLE	
Nominal system working pressure	90.0 [bar]
Nozzle k-value	0.62
Nozzle item number	SH11
Enclosure size	8.0 x 4.0 [m]
Tested ceiling height	2.5 [m]
Nozzle flow rate at 90.0 [bar]	5.88 [l/min]
Length of the supply hose (dry line)	60.0 [m]

Table of Contents

ABSTRACT.....	2
1. Background and Objective.....	4
2. Material Description	4
3. Test Description	5
3.1. Test Purpose	5
3.2. Test Facility	5
3.3. Test Configuration	6
3.4. Test procedure	10
4. Test Results	10
5. K- factor.....	11
6. Final conclusions	11
Appendix A – nozzle datasheet and drawings	12
Appendix B – test photos.....	13
Appendix C – test charts	17
Appendix D – cone calorimeter test certificate for foam	23
Appendix E – instrumentation list	28
Appendix F – laboratory certificates	81
TEST REPORT END.....	88

1. Background and Objective

Plumis initiated a test programme to assess the performance of its residential water mist system using the fire scenarios defined in BS EN 14972-17:2025.

As electronically actuated nozzles fall outside the scope of BS EN 14972-17:2025, formal compliance with Part 17 cannot be claimed for this system configuration.

To provide a relevant performance assessment, the Part 17 fire test scenarios were applied with a defined deviation whereby activation of the water mist nozzles was initiated by a fire detector.

The system utilised sidewall nozzles installed 1.2 m above the floor, spaced 4 m apart and positioned 2 m from the wall within an enclosure measuring 8 × 4 × 2.5 m (L × W × H). The objective of the testing was to verify suppression performance under the defined fire scenarios.

The water mist nozzles were received at Baltic Fire Laboratory, Tuchom, Poland on 17 November 2025. Non-accredited R&D testing was conducted on 25– 27 November 2025 at Baltic Fire Laboratory.

The purpose of the test was to evaluate the firefighting performance of an installed water mist system:

- SH11

The plywood sheets, sacrificial boards, wooden frames, polyether foam sheets BFL6F30, and wood crib sticks were conditioned at a temperature (23 ± 2)°C and a relative humidity of (50 ± 5)% in the calibrated climatic chamber BFL117. The wood cribs were conditioned before the tests in the calibrated thermostatic chamber BFL140 to a moisture content of (10 ± 2)%, measured 3 [mm] below the wood stick surface. Certificates are provided in Appendix D and Appendix E to this test report.

Six tests were performed accordance with BS EN 14972-17:2025. Three tests were carried out in a 8.0 x 4.0 [m] test enclosure with a ceiling height of 2.5 [m]. One test was performed in enclosed test enclosure with a ceiling height of 3.5 [m]. One test were performed in open space (corner test) with a ceiling height of 2.5 [m]. In all fire tests, the nozzles were installed on the wall at a height of 1.2 [m] above the floor.

The results in this report relate to the tested item (nozzle no. SH11) and configurations only.

2. Material Description

The water mist nozzles were supplied by Plumis. Specification of the nozzles is presented in Table 3.

Table 3. Water mist nozzle characteristic SH11.

WATER MIST NOZZLE	
Nominal system working pressure	90.0 [bar]
Nozzle k-value	0.62
Nozzle item number	SH11
Enclosure size	8.0 x 4.0 [m]
Tested ceiling height	2.5 [m]
Nozzle flow rate at 90.0 [bar]	5.88 [l/min]
Length of the supply hose	60.0 [m]

3. Test Description

3.1. Test Purpose

The purpose of test is to evaluate firefighting performance of Plumis high pressure system was to demonstrate the effectiveness of their water mist system in accordance with BS EN 14972-17:2025 „Fixed firefighting systems – Water mist systems – Part 17: Test protocol for residential occupancies for automatic nozzle systems”.

One type of the sidewall water mist nozzle has been subject to fire test performance: SH11.

3.2. Test Facility

Baltic Fire Laboratory is an international accredited fire testing laboratory for active fire protection systems full scale fire testing.

The fire test laboratory is accredited in accordance with PN EN ISO/IEC 17025:2018 by PCA accreditation Reg. AB1792.

In addition, Baltic Fire Laboratory is accepted as an independent 3rd part fire testing laboratory by IMO, USCG, UL, VdS.

Baltic Fire Laboratory (BFL) fire science complex consists of three buildings: laboratory office building, storage facility (conditioned building with controlled temperature and humidity) and fire laboratory hall in size 25.2 x 25.2 x 15.0 [m] (inside) equipped with the movable ceiling in dimensions 22.0 x 22.0 [m]. The movable ceiling allows to simulate ceiling in different heights of enclosures, to reproduce the height of the room in which the fire experiment (fire test) is performed.




Figure 1. Baltic Fire Laboratory, fire science complex.

The movable ceiling is comprised of steel grid and non-combustible wallboard. The surface of movable ceiling is additionally covered with 30 [mm] thick mineral wool board to insulate movable ceiling structure from heat created by fire experiment underneath.




Figure 2. Fire experiments laboratory hall with 9525 [m³] volume, movable ceiling highlighted.

Combustion air (fresh air) enters the building through 4 symmetrically placed louvers around the perimeter of the building. The exhaust fans (scrubber cleaning system) can be operated during fire test to allow for quick and efficient smoke extraction and maintain oxygen on required level by test method.

A continuous drain runs inside the perimeter of the building, leaving a 22.0 x 22.0 [m] inside inner concrete floor test area. The drain leads the waste liquid into an underground 105 [m³] fire - waste water tank.

Water was supplied from a 100 [m³] underground water tank via three centrifugal pumps located in laboratory pump room.

All fire test has been video recorded during all time of fire test. The control room also houses the data acquisition system, control systems for exhaust fans, and fire suppression system data.

Figure 1 & 2 shows a basic schematic of the test facility site.

3.3. Test Configuration

All tests were carried out in a 8.0 x 4.0 [m] test enclosure with a ceiling height of 2.5 [m] and one test in 3.5 [m]. In all fire tests, the nozzles were installed on the wall at a height of 1.2 m above the floor. The tests were conducted in accordance with BS EN 14972-17:2025. A total of six tests were performed. One corner test was conducted at height of 3.5 m and two corner tests were performed at height of 2.5 m. The remaining scenarios were performed only at a height of 2.5 m. These included one between two nozzles (B2) test, one corner test using a fan, and one corner test in an open enclosure. All tested fire scenarios are shown in Figure 3 below.

Table 4. Fire test series.

Fire test series	Fire tests for sidewall mounted nozzles	Test room ceiling height
Room ^a	F1	2.5 m and 3.5 m
Open space ^b	F2, F3, G	2.5 m or 3.5 m
	H	2.5 m or 3.5 m
^a Mandatory test series		
^b Optional test series		

Table 5. Description of fire tests for sidewall mounted nozzles.

Fire test	Description	Clause
F1, F2	Corner tests	7.3
F3	Between two nozzle test	7.3
G	Ventilation test	7.3
H	Open room test	7.3




Figure 3. Tested fire scenarios. The first row from the left represents corner tests in enclosure, with fan and open room test. The second row from the left includes one corner test with nozzles on the opposite side and one test conducted between two nozzles. The orange square indicates a wood crib placed on a tray. The blue dots represent nozzles N1, N2, and N3. The plywood is marked in brown. The green dots represent detectors in 2.5 [m] height enclosure. The pink dots represent detectors in 3.5 [m] height enclosure.




Figure 4. Thermocouples localisation.

Table 6. Thermocouple localisation.

Thermocouple no.	Localisation
T1	1.6 [m] above the floor
T2	75 [mm] below the ceiling
T3	1.6 [m] above the floor
T4	75 [mm] below the ceiling
T5	1.6 [m] above the floor
T6	75 [mm] below the ceiling
T7	75 [mm] below the ceiling
T8	1.6 [m] above the floor
T9	75 [mm] below the ceiling
T10	1.6 [m] above the floor
T11	75 [mm] below the ceiling
T12	1.6 [m] above the floor
T13	75 [mm] below the ceiling
T14	75 [mm] below the ceiling

3.4. Test procedure

The test begins with the simultaneous ignition of a tray filled with heptane, and cotton wicks. From the moment the first nozzle is activated, the test should continue for another 30 minutes.

Table 7. General test procedure.

Test Time (min:s)	Event
0:00	DAQ System/Cameras Rolling
1:00	Ignition of fuel sources
1:00~	Water pump activation, 1 st water mist nozzle activation
30:00 + Time of water pump activation	Shut off – water pump, test end

4. Test Results

After igniting the wood crib and wicks, the nozzle activation was awaited. From the activation of the first nozzle, the test was conducted for another 30 minutes.

Table 8. Tested fire scenarios.

Test no.	Scenario	Ceiling height [m]
251125-2	Corner test F1	3.5
251125-3	Corner test F1	2.5
251125-4	Corner F2	2.5
251125-5	Between two nozzles test (B2)	2.5
251126-1	Corner fan test	2.5
251127-1	Corner test – open room	2.5

Table 9. Summary of test records.

Test no.	251125-2	251125-3	251125-4	251125-5	251126-1	251127-1
Start	0:00	0:00	0:00	0:00	0:00	0:00
Fire ignition	1:00	1:00	1:00	1:00	1:00	1:00
Pump on	1:40	2:05	2:26	1:32	2:27	2:16
Nozzle activation	1:57	2:26	2:43	1:51	2:46	2:36
Test end	31:40	32:05	32:26	31:32	32:27	32:16

5. K- factor

K-factor measurement has been conducted for total 2 [pcs] nozzles using laboratory calibrated manometer, pressure transmitter and flow meter.

In the table 10 below, positive results considering the nozzle flow rate tolerance has been shown.

Table 10. Flow test nozzle SH11.

Sample	K-factor	Pressure [bar]	Nominal flow [l/min]	Flow measured indirect method [l/min]	Indirect method tolerance check $\pm 5\%$ (5.59 – 6.17 l/min)	FINAL COMMENT
N1	0.62	90.0	5.88 [l/min]	5.68	Within tolerance	Pass
N2	0.62	90.0	5.88 [l/min]	5.73	Within tolerance	Pass

6. Final conclusions

Six tests were performed accordance with BS EN 14972-17:2025. Three tests were carried out in a 8.0 x 4.0 [m] test enclosure with a ceiling height of 2.5 [m].

Total six fire tests has been conducted, one in 3.5 [m] and five in 2.5 [m] ceiling height enclosure.

In all fire tests, the nozzles were installed on the wall at a height of 1.2 [m] above the floor.

The tested water mist system met the criteria of BS EN 14972-17:2025.

The third nozzle, located outside the room, did not operate in any of the tests. The temperature values specified by the standard were not exceeded.

As electronically actuated nozzles fall outside the scope of BS EN 14972-17:2025, formal compliance with Part 17 cannot be claimed for this system configuration.


To provide a relevant performance assessment, the Part 17 fire test scenarios were applied with a defined deviation whereby activation of the water mist nozzles was initiated by a fire detector.

The system utilised sidewall nozzles installed 1.2 m above the floor, spaced 4 m apart and positioned 2 m from the wall within an enclosure measuring 8 × 4 × 2.5 m (L × W × H). The objective of the testing was to verify suppression performance under the defined fire scenarios.

A simple acceptance decision rule was chosen accordance with ILAC-G8:09/2019.

The results in this test report relate to the tested items and configurations only.

Appendix A – nozzle datasheet and drawings



*Figure A.1. Nozzle SH11– general assembly drawing.

*Information provided by the Customer.

Appendix B – test photos



Photo B.1. Test no. 251125-2, corner test, 3.5m – before water mist system activation.



Photo B.2. Test no. 251125-2, corner test, 3.5m – after water mist system activation.



Photo B.3. Test no. 251125-3, corner test, 2.5m – after water mist system activation.



Photo B.4. Test no. 251125-4, corner test, 2.5m – after water mist system activation.



Photo B.5. Test no. 251125-5, B2 test, 2.5m – after water mist system activation.




Photo B.6. Test no. 251126-1, corner fan test, 2.5m – after water mist system activation.



Photo B.7. Test no. 251127-1, corner test – open room, 2.5m – after water mist system activation.


Appendix C – test charts

Test no. 251125-2 - Temperatures [°C]




Graph C.1. Test no. 251125-2 – corner test, 3.5m, 30s average temperatures.

Test no. 251125-2 - Water flow and pressure




Graph C.2. Test no. 251125-2 – corner test, 3.5m, water flow and pressure.

Test no. 251125-3 - Temperatures [°C]




Graph C.3. Test no. 251125-3 – corner test, 2.5m, 30s average temperatures.

Test no. 251125-3 - Water flow and pressure




Graph C.4. Test no. 251125-3 – corner test, 2.5m, water flow and pressure.

Test no. 251125-4 - Temperatures [°C]




Graph C.5. Test no. 251125-4 – corner test, 2.5m, 30s average temperatures.

Test no. 251125-4 - Water flow and pressure




Graph C.6. Test no. 251125-4 – corner test, 2.5m, water flow and pressure.

Test no. 251125-5 - Temperatures [°C]




Graph C.7. Test no. 251125-5 – B2 test, 2.5m, 30s average temperatures.

Test no. 251125-5 - Water flow and pressure




Graph C.8. Test no. 251125-5 – B2 test, 2.5m, water flow and pressure.

Test no. 251126-1 - Temperatures [°C]




Graph C.9. Test no. 251126-1 – corner fan test, 2.5m, 30s average temperatures.

Test no. 251126-1 - Water flow and pressure




Graph C.10. Test no. 251126-1 – corner fan test, 2.5m, water flow and pressure.

Test no. 251127-1 - Temperatures [°C]



Graph C.11. Test no. 251127-1 – corner test – open room, 2.5m, 30s average temperatures.

Test no. 251127-1 - Water flow and pressure



Graph C.12. Test no. 251127-1 – corner test – open room, 2.5m, water flow and pressure.

Appendix D – cone calorimeter test certificate for foam



SYCHTA LABORATORIUM Sp. J.
Laboratorium Badań Palności Materialów
ul. Ofiar Stutthofu 90
72-010 Police



AB 1501

TEST REPORT

Order no: 01.12.2025

Signature: SL/Z-1316/ISO 5660-1/1549a/2025

Police, 15.12.2025

Test methods:

1. ISO 5660-1:2015/Amd.1:2019. Reaction to fire tests – Heat release, smoke production and mass loss rate –
Part 1: Heat release rate (cone calorimeter method)

Content of request: Reaction to fire test according to ISO 5660-1. Requirements:

Density (kg/m³): 27,2<...<30,4
HRR Peak (kW/m²): 345 +/- 85
Heat of combustion (MJ/kg): 22 +/- 3

Sponsor: Baltic Fire Laboratory Sp. z o.o.
St. Banińska 8
80-209 Tuchom
Poland

Material: Polyurethane foam - BFL6F30

Composition: Polyurethane foam with a density of 27,2-30,4 kg/m³
Thickness: 25 mm

Manufacturer/supplier: Baltic Fire Laboratory Sp. z o.o.
St. Banińska 8
80-209 Tuchom
Poland

Assessment: Density (kg/m ³):	30,0	27,2<...<30,4
HRR Peak (kW/m ²):	349	345 +/- 85
Heat of combustion (MJ/kg):	23,2	22 +/- 3

The reprint and the copying: only with the agreement of Baltic Fire Laboratory Sp. z o.o.

Without the written consent of the Sychta Laboratory the report can be copied only in one piece.

Report applies only to the sample tested and is not necessarily indicative of the qualities of apparently identical or similar products.

Content of test report: five pages with signature and numbers.

VAT PL8513152392
KRS 0000387681

phone +48 914210214
mobile +48 502078855

biuro@sychta.eu
www.sychta.eu



1. Heat release rate of specimen according to ISO 5660-1

Test conditions:

- irradiance: $30 \text{ kW} \cdot \text{m}^{-2}$
- horizontal position,
- sample thickness: $25,5 \pm 2 \text{ mm}$,
- no frame retainer shall be used.

Table 4. Heat release rate

Name of measured quantity	Unit	Specimen					Average	Standard deviation
		1	2	3	4	5		
Mass of the specimen	g	7,7	7,8	7,7	7,6	7,6	7,7	0,1
Specimen thickness	mm	25,4	25,6	25,6	25,6	25,6	25,6	0,1
Density	$\text{kg} \cdot \text{m}^{-3}$	30,5	30,3	29,9	29,8	29,7	30,0	0,3
Ignition time	s	2	4	2	2	2	2	1
Extinction time	s	48	58	62	74	54	59	10
Duration of the test	s	620	634	626	632	654	633	13
Maximum heat release rate	$\text{kW} \cdot \text{m}^{-2}$	342	370	340	360	332	349	16
Total heat release	$\text{MJ} \cdot \text{kg}^{-1}$	23,1	22,7	23,2	23,3	23,5	23,2	0,3
Maximum average rate of heat emission MARHE	$\text{kW} \cdot \text{m}^{-2}$	211,3	217,1	217,6	234,2	219,7	220,0	8,5

Remarks: none.




Figure 4.1. The relation of heat release rate and the time – specimen 1




Figure 4.2. The relation of heat release rate and the time – specimen 2




Figure 4.3. The relation of heat release rate and the time – specimen 3




Figure 4.4. The relation of heat release rate and the time – specimen 4




Figure 4.5. The relation of heat release rate and the time – specimen 5

2. Final findings

Density 27,2<...<30,4 (kg/m ³)	HRR Peak 345 +/- 85 (kW/m ²)	Heat of combustion 22 +/- 3 (MJ/kg)	Fulfilling the requirements
30,0	349	23,2	YES/NO

3. Remaining required information

Date of receipt of samples: 03.12.2025

Sampling: Sponsor took and delivered samples.

Description of the test material: Sponsor delivered 15 samples of white foam with description BFL6F30 with dimension of 100x100x25,3-25,7 mm and density of 29,7-30,5 kg/m³.

Conditioning of specimens: constant mass at a temperature of 23±2°C, and relative humidity of 50±5 %.

Declarations:

1. The test results relate to the behaviour of the test specimens under the particular conditions of the test; they are not intended to be the sole criterion for assessing the potential fire hazard of the products in use.
2. The information provided on the first page of the report concerning the scope of research and identification of the tested object/objects were provided by the Sponsor.

Operators:


mgr inż. Andrzej Sycha

SYCHTA LABORATORIUM Sp. J.
72-010 Police, ul. Ofiar Stutthofu 90
tel./fax +48 91 4210 214, tel. 502078855
e-mail: biuro@sycta.eu www:sycta.eu
KRS 0000387681 REGON 321023120
NIP 8513152392

Authorised by:


KIEROWNIK TECHNICZNY
dr inż. Krzysztof Sycha

Date and place of test - 10-11.12.2025, Police



Appendix E – instrumentation list

POS.	Description	Manufacturer	Model	Serial no.	Calibrated date	Cert no.
BFL13	Current meter	National Instruments	NI 9203	1CA668F	11.07.2024	2604/2024/E
BFL24	Anemometer	Testo	Testo 440	83508081	03.02.2025	42/A/25
BFL28	Moisture meter / Hygrometer / Thermometer (wood)	Testo	Testo 606-2	38782986/1021	23.08.2024	5011-KL-V0042-24
BFL45	Thermohygrometer	Meraserw	CEM DT-321S	210602678	07-11.03.2025	714/E/2025
BFL71	Pressure transmitter	Aplisens	PC-28	02245384	07.03.2024	0100/PCL/2024
BFL77	Manometer	Gesa	MD0906	2305252019852	10.05.2024	0637/2024/3
BFL83	Stopwatch	Casio	HS-80TW	K051188	12.06.2024	OUM6.6W2.473.14 35.2024.2
BFL85	Meter 5m	Stanley	-	1029/2024/1	12.08.2024	SW/1029/2024/1
BFL115	Flow meter	Endress Hauser	Promass F300	RA074602000	11.06.2025	20250611_F1_01
BFL117	Climatic chamber	-	-	-	05-06.06.2025	182/6/25
BFL120	Manometer	Gesa	MD0906	2206201176750	13.06.2025	1061/2025/3
BFL140	Thermostatic chamber	WAMED	SUP-200G	-	17.11.2025	848/11/25
BFL141	Balance	Radwag	WLC 30/F1/R	856846	11.10.2025	13817/4771/25/E
BFL1002	Thermocouple module	National Instruments	NI 9213	187F11C	31.10.2024	4100/2024/E
T1	PA2506021/10	Guenther	TYP K	1851220/46	24.06.2025	PA2506021
T2	PA2507022/4	Guenther	TYP K	1851220/24	14.07.2025	PA2507022
T3	PA2507023/1	Guenther	TYP K	1821220/17	14.07.2025	PA2507023
T4	PA2506021/6	Guenther	TYP K	1851220/31	24.06.2025	PA2506021
T5	PA2507023/3	Guenther	TYP K	1821220/37	14.07.2025	PA2507023
T6	PA2506021/9	Guenther	TYP K	1851220/39	24.06.2025	PA2506021
T7	PA2507022/1	Guenther	TYP K	1851220/13	14.07.2025	PA2507022
T8	PA2507023/2	Guenther	TYP K	1821220/20	14.07.2025	PA2507023
T9	PA2507022/6	Guenther	TYP K	1851220/37	14.07.2025	PA2507022
T10	PA2504047/2	Guenther	TYP K	1831220/44	14.04.2025	PA2504047/2
T11	PA2504112/4	Guenther	TYP K	1831220/18	30.04.2025	PA2504112
T12	PA2507022/5	Guenther	TYP K	1851220/36	14.07.2025	PA2507022
T13	PA2506021/7	Guenther	TYP K	1851220/33	24.06.2025	PA2506021
T14	PA2504112/3	Guenther	TYP K	1831220/17	30.04.2025	PA2504112

Baltic Fire Laboratory
BFL S/N: BFL13
Calib.Date: 11.07.2024
Next Calib.: 10.07.2026



Institute of Power Engineering – National Research Institute

MEASURING APPARATUS LABORATORY

02-981 Warszawa, ul. Augustówka 36

tel. (22) 34 51 438

lap@ien.com.pl

www.ien.com.pl



AP 013

Calibration laboratory accredited by
Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No AP 013

**CALIBRATION CERTIFICATE**

Date of issue: 12 July 2024

Certificate No: 2604/2024/E

Page 1/2

OBJECT OF CALIBRATION	Current meter model NI 9203, manufacturer NATIONAL INSTRUMENTS, serial number ICA668F, 8-channel with a current measuring chain.
APPLICANT	Baltic Fire Laboratory Sp. z o.o. 80 – 209 Tuchom, ul. Banińska 8 Order No: MAP.4032.864.2024
PLACE OF CALIBRATION	Baltic Fire Laboratory Sp. z o.o. 80 – 209 Tuchom, ul. Banińska 8
CALIBRATION METHOD	Calibration was performed according to the Measurement Procedure P4/MAP „Calibration apparatus for measuring temperature, pressure, electrical quantities and humidity on objects”, edition XII released on 30 May 2023.
ENVIRONMENTAL CONDITIONS	Ambient temperature (25,8 + 26,4) °C Relative humidity (66,2 + 73,6) %
DATE OF CALIBRATION	11 July 2024
TRACEABILITY	This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).
CALIBRATION RESULTS	The results have been presented on page 2/2 of this certificate including uncertainty of measurement. The result apply to the object of calibration only.
UNCERTAINTY OF MEASUREMENT	Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor $k = 2$.



Z-ca KIEROWNIKA
Laboratorium Aparatury Pomiarowej
[Signature]
Inż. Rafał Małkowski

This certificate may be presented or copied as a whole document only.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 013.

Certificate No: 2604/2024/E

Date of issue: 12 July 2024

CALIBRATION RESULTS

The results of the calibration are as follows:

REFERENCE VALUE W_p mA	AVERAGE MEASURED VALUE W_w mA								MEASUREMENT ERROR $W_w - W_p$ mA			MEASUREMENT UNCERTAINTY U mA
	1A	2A	3A	4A	5A	6A	7A	8A	1A	2A	3A	
	CH0	CH1	CH2	CH3	CH4	CH5	CH6	CH7	CH0	CH1	CH2	
Measuring chain →												
Mark on the card →												
Reading in the program → DAQ Express 4.1.0 Nr 7.1.0.50148												
4,000	3,993	3,992	3,990	4,006	4,000	3,994	3,998	4,002	-0,007	-0,008	-0,010	0,030
8,000	8,020	8,015	7,992	7,988	8,291	8,006	8,004	7,996	0,020	0,015	-0,008	0,030
12,000	11,990	12,005	11,998	12,012	11,995	12,012	11,995	12,000	-0,010	0,005	-0,002	0,030
16,000	16,012	16,000	15,985	16,004	16,004	15,980	15,997	15,979	0,012	0,000	-0,015	0,030
20,000	19,998	19,993	19,985	19,986	20,003	19,984	19,992	19,992	-0,002	-0,007	-0,015	0,030

Authorized by:

KIEROWNIK TECHNICZNY
Laboratorium Pomiarowe

Inż. Rafał Malinowski



**LABORATORY FOR CALIBRATION
OF THE VENTILATION MEASURING DEVICES**
STRATA MECHANICS RESEARCH INSTITUTE of POLISH ACADEMY OF SCI.
30-059 CRACOW, POLAND Reymonta 27 Str. phone.: +48 12 6376200; fax: +48 12 6372884
www.hwppp.imgpan.pl; hwppp@imgpan.pl



Laboratory accredited by
Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No AP 118.

CALIBRATION CERTIFICATE

Date of issue: February 3 2025

Certificate No: 42/A/25

Page 1/2

**OBJECT OF
CALIBRATION**

Name: Vane anemometer
Type: Testo 440, probe 0635 9370
Factory No: 83508081, probe no. 20980936
Range: (0,1 ÷ 15,0) m/s
Reading accuracy: 0,01 m/s
Manufacturer: Testo

Baltic Fire Laboratory
BFL S/N: BFL24
Calib.Date: 03.02.2025
Next Calib.: 03.02.2027

APPLICANT

Baltic Fire Laboratory Sp. z o.o.
80-209 Tuchom, ul. Banińska 8

Baltic Fire Laboratory
BFL S/N: BFL42
Calib.Date: 03.02.2025
Next Calib.: 03.02.2027

**CALIBRATION
SITE**

Laboratory for Calibration of the Ventilation Measuring Devices
30-059 Cracow, Poland, Reymonta 27 Str.

**CALIBRATION
METHOD**

Anemometers calibration instruction no P/01 of 15.07.2022 r.

**ENVIRONMENTAL
CONDITIONS**

Temperature: (19,90 ÷ 20,14) deg C
Barometric pressure: (1003,47 ÷ 1003,69) hPa
Air density: (1,18 ÷ 1,18) kg/m³

**DATE OF
CALIBRATION**

February 3 2025

TRACEABILITY

This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).

**CALIBRATION
RESULTS**

Printed on the second page and they concern only the calibration object.

**UNCERTAINTY
OF MEASUREMENT**

Measurement uncertainty has been evaluated in compliance with the EA-4/02 M:2022.
The expanded uncertainty assigned corresponds to a coverage probability of 95% and the coverage factor k = 2.

Acting Head of the Calibration Laboratory

Dr Eng. Jakub Janus

This certificate may be presented or copied as a complete document only.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 118

Date of issue: February 3 2025

Certificate No: 42/A/25

Page 2/2

CALIBRATION RESULTS

Calibration results are the following:

Lp.	Reference Velocity		Velocity indicated by anemometer		Correction	Measurement uncertainty
	V		Va		DV=V-Va	U(Dv)
	m/s	m/min	m/s	m/min	m/s	m/s
1	0,15	9	0,12	7	0,03	0,05
2	1,00	60	1,04	62	-0,04	0,05
3	3,00	180	3,06	184	-0,06	0,14
4	5,00	300	5,15	309	-0,15	0,11
5	10,00	600	10,30	618	-0,30	0,11
6	15,00	900	15,27	916	-0,27	0,16

Checked by


 Dr Eng. Paweł Jamróz

**Český metrologický institut**

Okružní 31, 638 00 Brno

tel. +420 545 555 111

www.cmi.cz

**K 2202****Calibration laboratory No. 2202 accredited by the Czech Accreditation Institute according to ISO/IEC 17025:2017****Laboratory:** Regional Inspectorate Pardubice, Průmyslová 455, 530 03 Pardubice
Department of moisture of solids, tel.: 466 670 696; e-mail : lpitrovanetolicka@cmi.cz

CERTIFICATE OF CALIBRATION

5011-KL-V0042-24**Date of issue:** August 23th, 2024

Page 1 of 2

Customer: Baltic Fire Laboratory
ul. Banišská 8
80-209 Tuchom

Baltic Fire Laboratory
BFL S/N: BFL28
Calib.Date: 23.08.2024
Next Calib.: 22.08.2026

User*: Baltic Fire Laboratory
ul. Banišská 8
80-209 Tuchom**Measuring instrument:** Moisture meter of wood**Manufacturer:** TESTO**Type:** Testo 606-2**Serial number:** 38782986/1021**Identific. number:** BFL 28**Measuring range:** (7 – 78) % - for material curve 2**Number of calibration scales:** 1 - pine

The results of the calibration have been obtained following the procedures reported in this Certificate and are related only to the calibrated measuring instrument, the date, place and conditions of the calibration.

Date of calibration: August 22-23th, 2024**Calibrated by:****Director of the regional inspectorate:**

Ing. Lucie Pitrová Netolická



Ing. Zuzana Pálková, Ph.D.

This document may only be reproduced in full, except with the prior written permission by the issuing laboratory.

CERTIFICATE OF CALIBRATION
5011-KL-V0042-24

Page 2 of 2

Metrological traceability:	During the calibration of the moisture meter was used secondary moisture standard of solid substances – wood, SEVPL-D, serial number: 001/92, range $(4 \div 110) \% [m/m]$ and digital thermometer - hygrometer, reference number: 970053. The measurements are metrologically traceable to national/international standards.
Calibration procedure:	Tests of the meter were performed according to calibration procedure No. 511-MP-C003 Calibration of moisture meters for wood. The moisture content of the wood samples used for calibration was adjusted by conditioning. The measurement was carried out according to the working manual for the tested hygrometer. Type of material used for calibration: wood - pine Calibration curve set on a calibrated gauge: 2 (pine)
Place of calibration: (complete address)	ČMI Ol Pardubice - laboratoř VPL, Průmyslová 455, 530 03 Pardubice
Ambient conditions: (nominal values)	1) temperature: $(23 \pm 5)^\circ\text{C}$ 2) relative humidity: max. 70 %

Results of calibration:

No. sample	BO11	BO12	BO13	BO21	BO22	BO23	BO31	BO32	BO33
w_h	7.30	7.65	7.65	10.80	10.93	10.55	17.73	18.05	18.48
w_r	8.81	8.60	8.65	11.82	12.34	12.40	18.76	19.48	19.07
s_r	-1.51	-0.95	-1.00	-1.02	-1.42	-1.85	-1.04	-1.43	-0.59
U_c	0.42	0.54	0.87	0.89	0.53	0.95	0.95	0.91	0.91

Explanations:

- line 2: w_h : mean of 4 moisture measurements of the sample with the tested moisture meter in % moisture
 line 3: w_r : reference gravimetric moisture determination of the sample in % moisture
 line 4: error of moisture meter $s_r = w_h - w_r$ in % humidity
 line 5: uncertainty U_c in % moisture for individual samples

The standard uncertainty of measurement has been determined in accordance with EA-4/02 M:2022 document.
 The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k corresponding to a coverage probability of approximately 95 %, which for normal distribution corresponds to a coverage factor $k = 2$.

* ... data from customer

End of calibration certificate.

 Český metrologický institut
 Oblastní inspektorát Pardubice
 Průmyslová 455
 530 03 Pardubice
 26...

This document may only be reproduced in full, except with the prior written permission by the issuing laboratory.



Institute of Power Engineering – National Research Institute
MEASURING APPARATUS LABORATORY
02-981 Warszawa, ul. Augustówka 36
tel. (22) 34 51 438
lap@ien.com.pl www.ien.com.pl



AP 013

Baltic Fire Laboratory
BFL S/N: BFL45
Calib.Date: 07-11.03.2025
Next Calib.: 11.03.2027

Calibration laboratory accredited by
Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates,
Accreditation No AP 013



CALIBRATION CERTIFICATE

Date of issue: 11 March, 2025

Certificate No: 714/E/2025

Page 1/2

OBJECT OF CALIBRATION	Termohygrometer model DT-321S, manufacturer CEM, serial number 210602678, inventory number BFL45 with integrated external sensor.
APPLICANT	Baltic Fire Laboratory Sp. z o.o. 80 - 209 Tuchom, ul. Banińska 8 Order No: MAP.4032.245.2025
CALIBRATION METHOD	Calibration was performed at the premises of the LAP IEN-PIB according to the Measurement Procedure W1/MAP "Calibration of Humidity Measuring Instruments", XII edition released on 18 September 2020.
ENVIRONMENTAL CONDITIONS	Ambient temperature (22,6 + 26,8) °C Relative humidity (25,6 + 37,9) %
DATE OF CALIBRATION	7 March ÷ 11 March 2025
TRACEABILITY	This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).
CALIBRATION RESULTS	The results have been presented on page 2/2 of this certificate including uncertainty of measurement. The results apply to the object of calibration only.
UNCERTAINTY OF MEASUREMENT	Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor $k = 2$.



Z-ca KIEROWNIKA
Laboratorium Aparatury Technicznej
[Signature]
inż. Rafał Malinowski

This certificate may be presented or copied as a whole document only.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 013.

Date of issue: 11 March, 2025

Certificate No: 714/E/2025

Page 2/2

CALIBRATION RESULTS

Calibration results are the following:

AVERAGE REFERENCE VALUE W_p		AVERAGE MEASURED VALUE W_w		MEASUREMENT ERROR $W_w - W_p$		MEASUREMENT UNCERTAINTY U	
t	RH	t	RH	t	RH	t	RH
°C	%	°C	%	°C	%	°C	%
10,08	49,5	10,09	49,3	0,01	-0,2	0,15	0,7
19,98	29,6	19,99	27,3	0,01	-2,3	0,15	0,6
19,95	49,5	19,96	44,6	0,01	-4,9	0,15	0,7
19,99	69,5	20,02	64,0	0,03	-5,5	0,15	1,0
29,92	50,2	29,91	46,6	-0,01	-3,6	0,15	0,7
49,94	50,3	49,91	47,1	-0,03	-3,2	0,15	0,8

Temperatures given are referred to the International Temperature Scale of 1990 (ITS-90).

Authorized by:

KIEROWNIK TECHNICZNY
 Laboratorium Aparatury Pomiarowej

 inż. Rafał Malinowski

		APLISENS S.A. LABORATORIUM WZORCZĄCE APLISENS 03-192 Warszawa, ul. Morelowa 7 tel.: 22 814 07 77 w. 162, faks: 22 814 07 78 e-mail: laboratorium@aplisens.pl http://www.aplisens.pl PRACOWNIA CIŚNIEŃ 03-192 Warszawa, ul. Morelowa 7 e-mail: laboratorium@aplisens.pl		 AP 122	
Baltic Fire Laboratory BFL S/N: BFL71 Calib.Date: 07.03.2024 Next Calib.: 06.03.2026		Laboratorium wzorczące akredytowane przez Polskie Centrum Akredytacji, sygnatariusza porozumień EA MLA i ILAC MRA dotyczących wzajemnego uznawania świadectw wzorcowania. Nr akredytacji AP 122. Calibration laboratory accredited by Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA that include recognition of calibration certificates. Accreditation No. AP 122. ŚWIADECTWO WZORCOWANIA CALIBRATION CERTIFICATE 			
Data wydania: Date of issue	07.03.2024	Numer świadectwa: Certificate No.	0100/PCU/2024	Strona nr: 1/2 Page	
OBIEKT WZORCOWANIA OBJECT OF CALIBRATION	Nazwa: Name	Przetwornik ciśnienia Pressure transmitter	Nr fabryczny: Serial Number	02245384	
	Typ: Type	PC-28	Producent: Producer	APLISENS	
	Zakres wskazań: Indication range	(0÷25) MPa			
	Sygnał wyjściowy: Output signal	(4÷20) mA, charakterystyka liniowa (4÷20) mA, with linear characteristic			
ZGŁASZAJĄCY APPLICANT	Baltic Fire Laboratory Sp. z o.o. ul. Banińska 8 80-209 Tuchom				
MIEJSCE WZORCOWANIA PLACE OF CALIBRATION	APLISENS S.A., Laboratorium Wzorczące Aplsens - Pracownia Ciśnień ul. Morelowa 7, 03-192 Warszawa				
METODA WZORCOWANIA CALIBRATION METHOD	Zgodnie z Procedurą PR-01 "Wzorcowanie przetworników ciśnienia" wydanie B z dnia 14.11.2023 r According to PR-01 Procedure: Calibration of pressure transmitters, edition B from 14.11.2023				
WARUNKI ŚRODOWISKOWE ENVIRONMENTAL CONDITIONS	Temperatura otoczenia Environment temperature	Wilgotność względna Relative humidity	Ciśnienie atmosferyczne Atmospheric pressure		
	(19,7÷20,4) °C	(27,9÷31,9) %	(1000,0÷1016,6) hPa		
DATA WZORCOWANIA DATE OF CALIBRATION	07.03.2024				
SPÓJNOŚĆ POMIAROWA TRACEABILITY	Świadectwo jest wydane w ramach porozumienia EA MLA w zakresie wzorcowania i potwierdza spójność wyników pomiarów z jednostkami miar Międzynarodowego Układu Jednostek Miar (SI). This certificate is issued under the agreement EA MLA in the calibration field and provides traceability of measurement results to the International System of Units (SI).				
WYNIKI WZORCOWANIA CALIBRATION RESULTS	Podano na stronie 2 niniejszego świadectwa wraz z wartościami niepewności pomiaru i dotyczą wyłącznie obiektu wzorcowanego. The results have been presented on 2 page of this certificate, including the measurement uncertainty, and they apply to the calibrated object only.				
NIEPEWNOŚĆ POMIARU UNCERTAINTY OF MEASUREMENT	Niepewność pomiaru została określona zgodnie z dokumentem EA-4/02 M:2022. Podane wartości niepewności stanowią niepewności rozszerzone przy prawdopodobieństwie rozszerzenia ok. 95% i współczynniku rozszerzenia k=2. Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor k=2.				
LABORATORIUM WZORCZĄCE APLISENS / PRACOWNIA CIŚNIEŃ					
 ZSW259986180KCL000123PC24000093 					
Niniejsze świadectwo może być okazywane lub kopiowane tylko w całości. This certificate may be presented or copied as a whole document only.					

ŚWIADECTWO WZORCOWANIA wydane przez Laboratorium Wzorcujące Aplisens Nr AP 122 CALIBRATION CERTIFICATE issued by Laboratorium Wzorcujące Aplisens No. AP 122							
Data wydania: Date of issue	07.03.2024	Numer świadectwa: Certificate No.	0100/PCU/2024	Strona nr: 2/2 Page			
WYNIKI WZORCOWANIA CALIBRATION RESULTS							
Wyniki wzorcowania obiektu przedstawiono poniżej: Calibration results are the following:							
Wartość ciśnienia odniesienia <i>Reference pressure value</i>	Wartość natężenia prądu odpowiadająca wartości ciśnienia odniesienia <i>Current strenght value related to reference pressure value</i>	Zmierzona wartość natężenia prądu <i>Measured current strenght value</i>	Wartość ciśnienia odpowiadająca zmierzonej wartości natężenia prądu <i>Pressure value related to measured current strenght value</i>	Błąd pomiaru <i>Measurement error</i>		Niepewność pomiaru <i>Measurement uncertainty</i>	
MPa	mA	mA	MPa	mA	MPa	mA	MPa
0,0000	4,000	4,000	0,0000	0,000	0,0000	0,001	0,0016
6,2500	8,000	7,997	6,2453	-0,003	-0,0047	0,001	0,0016
12,5001	12,000	11,995	12,4822	-0,005	-0,0079	0,002	0,0031
18,7503	16,000	15,995	18,7422	-0,005	-0,0081	0,003	0,0047
25,0005	20,000	19,998	24,9969	-0,002	-0,0036	0,004	0,0062
KOMENTARZ COMMENT							
Podczas wzorcowania przetwornik umieszczono w pozycji pionowej czujnikiem w dół. <i>The transmitter was placed in vertical position with the sensor down during the calibration.</i>							
mgr inż. Barbara Bucek  <i>Kierownik Laboratorium Wzorcujące Aplisens</i> 2024-03-07 Autoryzował(a): Authorized by:							
Niniejsze świadectwo może być okazywane lub kopiowane tylko w całości. <i>This certificate may be presented or copied as a whole document only.</i>							

Baltic Fire Laboratory
BFL S/N: BFL77
Calib.Date: 10.05.2024
Next Calib.: 09.05.2026



TENSLAB Spółka z ograniczoną odpowiedzialnością
LABORATORIUM WZORCUJĄCE

Siedziba: 80-554 GDAŃSK
ul. Śnieżna 5

Laboratorium: 81-336 GDYNIA
ul. Kadłubowców 4

Kontakt: www.tenslab.pl
tel: +48 58 380 06 44



AP 180

Calibration laboratory accredited by Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA that include recognition of calibration certificates.
Accreditation No AP 180.



CALIBRATION CERTIFICATE

Date of issue: 10 May 2024

Certificate No: 0637/2024/3

Page 1 / 2

OBJECT OF CALIBRATION	Digital pressure gauge, Type: MD0906, Serial number: 2305252019852, Producer: Gesa, Resolution: 0,01 bar, Measuring range: 0,0+250,0 bar, Class: 0,2, Identification number: no data.
CUSTOMER	BALTIC FIRE LABORATORY SPÓŁKA Z O. O. ul. Banińska 8 80-209 Tuchom
PLACE OF CALIBRATION	TENSLAB Sp. z o.o. Laboratorium Wzorcujące ul. Kadłubowców 4 81-336 Gdynia
CALIBRATION METHOD	Procedura pomiarowa - Ciśnieniomierze elektroniczne - PP TENSLAB/CE/1/18 date of issue 20.07.2018
ENVIRONMENTAL CONDITIONS	Ambient temperature: (21,7 + 22,3) °C
DATE OF CALIBRATION	10 May 2024
TRACEABILITY	This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).
CALIBRATION RESULTS	The results have been presented on page(s) 2 of this certificate including uncertainty of measurement. The results relate only to the object of calibration.
UNCERTAINTY OF MEASUREMENT	Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor k = 2.



This Calibration Certificate can be shown or copy only as a whole.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 180

Date of issue: 10 May 2024

Certificate No: 0637/2024/3

Page 2/ 2

**CALIBRATION
RESULTS**

Calibration results are the following:

Determination of measurement error:

No	Measured pressure value	Reference pressure value	Measurement error	Hysteresis	Measurement uncertainty
	bar				
1.	0,00	0,00	0,00	0,00	0,10
2.	5,00	5,01	-0,01	0,04	0,10
3.	16,00	15,97	0,03	0,07	0,10
4.	25,00	24,90	0,10	0,14	0,12
5.	50,00	49,79	0,21	0,16	0,13
6.	75,00	74,76	0,24	0,16	0,13
7.	100,00	99,75	0,25	0,15	0,12
8.	150,00	149,75	0,25	0,19	0,14
9.	200,00	199,70	0,30	0,15	0,13
10.	225,00	224,72	0,28	0,10	0,12
11.	250,00	249,76	0,24	0,10	0,11
12.	-	-	-	-	-
13.	-	-	-	-	-
14.	-	-	-	-	-
15.	-	-	-	-	-

Authorized by:

MEROWNIK TECHNICZNY


Paula Koister-Lyko

END OF CALIBRATION CERTIFICATE

					
Baltic Fire Laboratory BFL S/N: BFL83 Calib.Date: 12.06.2024 Next Calib.: 11.06.2026		DIRECTOR OF THE REGIONAL OFFICE OF MEASURES IN GDAŃSK			
Calibration Laboratories Group Laboratory of Electricity and Physicochemistry Polanki 124c Str., 80-308 Gdańsk tel.: 58-5245-200 fax: 58-5245-229 e-mail: w2@oum.gda.pl www.oum.gda.pl					
CALIBRATION CERTIFICATE					
Date of issue: 12 June 2024		Certificate No: OUM6.6W2.473.1435.2024.2.EN		Page 1/2	
OBJECT OF CALIBRATION	Electronic stopwatch Manufacturer: Casio Indicate range: 9 h 59 min 59,999 s	Type: No: Resolution:	HS-80TW K051188 0,001 s		
APPLICANT	Gdańskie Usługi Metrologiczne Sp. z o.o. ul. Kręta 18 80-217 Gdańsk				
USER	Baltic Fire Laboratory Sp. z o.o. ul. Banińska 8 80-209 Tuchom				
PLACE OF CALIBRATION	At the Laboratory of Electricity and Physicochemistry facility				
CALIBRATION METHOD	"Wzorcowanie stoperów elektronicznych wyzwalanych ręcznie" PP/6W2/524/02 wydanie 6 z dnia 16 września 2020 r.				
ENVIRONMENTAL CONDITIONS	Temperature: (23,8 ± 24,4) °C				
DATE OF CALIBRATION	12 June 2024				
TRACEABILITY	This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).				
CALIBRATION RESULTS	The results apply to the calibrated object only and have been presented on page 2/2 of this certificate including uncertainty of measurement. niepewności pomiaru.				
UNCERTAINTY OF MEASUREMENT	Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor k= 2.				
					
AP 086				z up. Dyrektora Okręgowego Urzędu Miar w Gdańsku <i>Małgorzata Siedziowska</i> Główny Specjalista Laboratorium Elektryczności i Fizykochemii	
				F2/P-11/10.02.2023	
					

This certificate may be presented or copied only in its entirety.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 086

Date of issue: 12 June 2024

Certificate No: OUM6.6W2.473.1435.2024.2.EN

Page 2/2

**CALIBRATION
RESULTS**

Calibration results are the following:

1) form time interval $\tau = 1h$

error for 1h time interval	uncertainty
s	s
0,011	0,010

2) of any time interval τ :

$$\Delta \tau = y \tau \pm 0,010 \text{ s}$$

y - relative stopwatch error

$\Delta \tau$ - measurement error

y in time calibration:

y	uncertainty
$\times 10^{-6}$	$\times 10^{-6}$
3,01	0,30

Notice

The stated measurement error does not take into account the subjective error associated with the reaction time of the person taking the measurements.

Authorized by:

STARSZY INSPEKTOR


mgr Ewa Kozienowska

THE END OF THE CALIBRATION CERTIFICATE



Baltic Fire Laboratory
BFL S/N: BFL85
Calib.Date: 12.08.2024
Next Calib.: 12.08.2028



"LAB MESK" EUGENIUSZ STEPNOWSKI
01-441 Warsaw, Deotymy Street 43a apt.6

Calibration Laboratory Lab Mesk

Wólczyńska Street 133, building no. 5, apt no. 541, 01-919 Warsaw
tel.: 505-414-661, 500-002-611 | labmesk.pl | e-mail: laboratorium@labmesk.pl

Provided address is permanent place of laboratory activities

Calibration laboratory accredited by
Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No AP 171.



AP 171



CALIBRATION CERTIFICATE

Date of issue: 12th August 2024 Certificate No: SW/1029/2024/1 Page: 1/2

OBJECT OF CALIBRATION	Measuring tape	
	Manufacturer:	Stanley
	Manufacturing number:	-
	Identification number:	1029/2024/1
	Range:	5,0 m
	Graduation:	1 mm
APPLICANT	Baltic Fire Laboratory Sp. z o.o. Banińska 8, 80-209 Tuchom	
CALIBRATION METHOD	Calibration procedure "IW5.8.21 Przymiary" version 2 of 20th September 2022	
ENVIRONMENTAL CONDITIONS	Temperature:	(19,3 ÷ 19,7)°C
DATE OF CALIBRATION	12th August 2024	
TRACEABILITY	This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).	
CALIBRATION RESULTS	The results have been presented on page 2 of this certificate including uncertainty of measurement The results refer only to the instrument mentioned in the field "object of calibration".	
UNCERTAINTY OF MEASUREMENT	Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor $k = 2$.	



KIRBOWNIK
LABORATORIUM
Marek Yoka

This certificate may be presented or copied as a whole document only.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No. AP 171

Date of issue: 12th August 2024

Certificate No: SW/1029/2024/1

Page: 2/2

**CALIBRATION
RESULTS**

Calibration results are the following:

Nominal length of the calibrated section	Permissible length of the calibrated section	Measurement uncertainty
mm	mm	mm
Upper scale		
0÷100 docisk	100,13	0,14
0÷100 zaczep	100,08	
0÷1000 zaczep	1000,02	
0÷2000 zaczep	2000,30	
0÷3000 zaczep	3000,36	0,25
0÷4000 zaczep	4000,44	
0÷5000 zaczep	5000,78	
Lower scale		
0÷100 docisk	100,13	0,14
0÷100 zaczep	100,08	
0÷1000 zaczep	1000,02	
0÷2000 zaczep	2000,30	
0÷3000 zaczep	3000,36	0,25
0÷4000 zaczep	4000,44	
0÷5000 zaczep	5000,78	

Authorized by:

 KIEROWNIK
 LABORATORIA
 Martin Jozak

End of calibration certificate.

N.o.: 20250611_F1_01


Endress+Hauser 
People for Process Automation

Accredited Calibration Laboratory according to: NAH 2-0327/2021

Calibration certificate

Baltic Fire Laboratory
BFL S/N: BFL115
Calib.Date: 11.06.2025
Next Calib.: 11.06.2027

Subject:	Promass F 300 Calibration		
Manufacturer:	Endress+Hauser		
Type / Order code:	Promass 8F3B15-D7C4/0		
Serial number: / Tag number:	no. RA074602000 - BFL 115		
Measuring range / calibrated range	0-108,3 kg/min / 1,66-108,3 kg/min		
Output signal:	4-20 mA		
Calibration factor:	1.9828		
Ordered by:	Baltic Fire Laboratory		
Place and date of calibration:	E+H Cal. Lab. 11.06.2025		
Calibrated by:	Lukács László		
Status:	OK for calibration		
References	Order code	Serial n.:	Certificate
Promass 83F 83F08-AD2SAMABAAA	J	9B098002000	N11203636
Promass 83F 83F25-AD2SAMABAAA	J	9B097F02000	N11203740
Promass 83F 83F50-AD2SAMABAAA	J	CB0A2C02000	N11203392
TR11-ARACARY17000		9C005E142FD	WT1804393
TTR31-A1A111YY2CAB		F2004C0426E	WT1504267

The measurement results were traced to the national standard.

The calibration certificate contains data specific to the time and circumstances of the calibration of the instrument.

Calibration method:

The calibration was performed according to KE-01/2019.01.10 ver.:1

Conditions of calibration:

Water temperature: 21,3 °C
Ambient temperature: 21,9 °C
Measuring fluid: tapwater

Stamps:

The calibration is certified by the 20250611_F1_01 stamp on the device.

This certificate can only be copied in its full form and scope

Endress+Hauser Magyarország
Folyamatműszerezési Kft.
1139 Budapest
Forgách u. 9/b.

1/2

Telefon +36 1 412 0421
Telefax +36 1 412 0424
hungary@hu.endress.com
www.hu.endress.com

N.o.: 20250611_F1_01


Endress+Hauser 

People for Process Automation

Accredited Calibration Laboratory according to: NAH 2-0327/2021

Results:


Set points	Set points	Reference value	Measured value	Measuring error	Measurement Uncertainty
%	kg/min	kg	kg	%	%
1,55%	1,676	1,677	1,680	0,18%	0,15
23,08%	25,003	25,003	25,011	0,03%	0,15
46,19%	50,040	50,041	50,030	-0,02%	0,15
70,30%	76,152	75,518	75,496	-0,03%	0,15
99,21%	107,473	108,371	108,357	-0,01%	0,15

Measurement uncertainty:

The uncertainty values given in the tabloids are extended ones which include uncertainties coming from the etalons, from the method of calibration and from enviromental conditions. In case of a normal (Gauss) distribution, the value multiplied by $k=2$ fits a 95% cover probability. The standard uncertainty was determined by the publication no. EA 4/02M of the European Accrediting Organization.


Validity of the certificate:

The date of a repeated inspection is to be determined by the user with regard to the operation circumstances of the measurement device



Calibrated by
Laszlo Lukacs

ENDRESS+HAUSER
MAGYARORSZÁG KFT.
1139 Budapest, Forgách u. 9./b.
Tel: 412-0421 • Fax: 412-0424
www.hu.endress.com • E-mail: hungary@hu.endress.com



Checked by
Laszlo Nemes

This certificate can only be copied in its full form and scope

Endress+Hauser Magyarország
Folyamatműszerezési Kft.
1139 Budapest
Forgách u. 9./b.

Telefon +36 1 412 0421
Telefax +36 1 412 0424
hungary@hu.endress.com
www.hu.endress.com



LABORATORIUM USŁUGOWO-BADAWCZE

„BIOCHEMIK” Sp. z o.o.

Pracownia Wzorcowania

ul. Strefowa 15, 64 – 920 Piła

Mob. +48 727 630 955, e-mail: wzorcowanie@biochemik.pl, www.wzorcowanie.biochemik.pl

Calibration laboratory accredited by

Polish Centre of Accreditation, a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.

Accreditation No AP 158



AP 158



CALIBRATION CERTIFICATE

Date of issue: 12th June 2025

Certificate No: 182/6/25

Page 1/4

OBJECT OF CALIBRATION

Climatic chamber

Manufacturer: -

Type: -

Serial number: -

User marking: BFL 117

Resolution: 0,1 °C, 0,1 %

Workspace dimensions: (width x depth x height): (2280 x 5400 x 2280) mm

Baltic Fire Laboratory
BFL S/N: BFL117
Calib.Date: 05-06.06.2025
Next Calib.: 05.06.2027

APPLICANT

Baltic Fire Laboratory Sp. z o.o.

ul. Banińska 8

80-209 Tuchom, Poland

PLACE OF CALIBRATION

Baltic Fire Laboratory Sp. z o.o.

ul. Banińska 8

80-209 Tuchom, Poland

CALIBRATION METHOD

Calibration has been carried out according to the calibration procedure „Climatic and thermostatic chambers”, PW-06, 7th edition of 02 January 2022.

ENVIRONMENTAL CONDITIONS

Temperature: (16,8 ÷ 19,8) °C

Relative humidity: (65,4 ÷ 74,6) %

DATE OF CALIBRATION

5th June ÷ 6th June 2025

TRACEABILITY

This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).

CALIBRATION RESULTS

The results have been presented on page 2/4 of this certificate including uncertainty of measurement. The results apply only to the calibration object.

UNCERTAINTY OF MEASUREMENT

Uncertainty of measurement has been evaluated in compliance EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95% and the coverage factor $k=2$.



Head of the Calibration
Laboratory



M.Sc.Eng. Paweł Janik

This certificate may be presented or copied as a whole document only.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 158

Date of issue: 12th June 2025

Certificate No: 182/6/25

Page 2/4

CALIBRATION RESULTS

Calibration results are the following:

Set value of the test object		Average measured reference value		Average temperature indication of the test object		Measurement error		Measurement uncertainty ¹⁾	
<i>t</i>	<i>RH</i>	<i>t_w</i>	<i>RH_w</i>	<i>t_x</i>	<i>RH_x</i>	Δt	ΔRH	<i>U</i>	
°C	%	°C	%	°C	%	°C	%	°C	%
Measuring location L1, Point P2									
-18,0	-	-18,6	-	-18,6	-	0,0	-	2,2	-
19,0	90,0	19,2	92,5	19,0	92,6	-0,2	0,1	0,2	3,5
21,0	90,0	21,2	90,5	21,0	91,6	-0,2	1,1	0,3	6,4
25,0	80,0	25,1	81,3	25,0	82,0	-0,1	0,7	0,6	3,5

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

¹⁾The measurement uncertainty include into account the spatial inhomogeneity of the chamber.

TEMPERATURE DISTRIBUTION IN THE CHAMBER (at temperature setting: -18,0 °C)

Measuring point	Average value of reference temperature at the measuring point	Minimum value of the reference temperature at the measuring point	Maximum value of the reference temperature at the measuring point	Measurement uncertainty ³⁾
	°C	°C	°C	°C
L1P1	-18,83	-20,11	-17,75	2,19
L1P2	-18,55	-19,65	-17,62	2,06

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾The measurement uncertainty does not include into account the spatial inhomogeneity of the chamber.

TEMPERATURE AND RELATIVE HUMIDITY DISTRIBUTION IN THE CHAMBER

(at humidity setting: 90,0 % at 19,0 °C)

Measuring point	Average value of reference humidity and temperature at the measuring point		Minimum value of reference humidity and temperature at the measuring point		Maximum value of reference humidity and temperature at the measuring point		Measurement uncertainty ³⁾	
	°C	% RH	°C	% RH	°C	% RH	°C	% RH
L1P1	19,21	93,85	19,12	93,20	19,28	94,70	0,20	2,36
L1P2	19,20	92,51	19,12	91,76	19,28	93,36	0,20	3,11

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾The measurement uncertainty does not include into account the spatial inhomogeneity of the chamber.

Authorized by:
M.Sc.Eng. Marta Piotrowska

Additional information:

Document signed with a qualified electronic signature.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 158

Date of issue: 12th June 2025

Certificate No: 182/6/25

Page 3/4

TEMPERATURE AND RELATIVE HUMIDITY DISTRIBUTION IN THE CHAMBER

(at humidity setting: 90,0 % at 21,0 °C)

Measuring point	Average value of reference humidity and temperature at the measuring point		Minimum value of reference humidity and temperature at the measuring point		Maximum value of reference humidity and temperature at the measuring point		Measurement uncertainty ²⁾	
	°C	% RH	°C	% RH	°C	% RH	°C	% RH
L1P1	21,24	91,90	21,16	87,87	21,42	93,27	0,25	5,94
L1P2	21,23	90,52	21,15	86,54	21,39	91,94	0,23	6,23

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

²⁾The measurement uncertainty does not include into account the spatial inhomogeneity of the chamber.

TEMPERATURE AND RELATIVE HUMIDITY DISTRIBUTION IN THE CHAMBER

(at humidity setting: 80,0 % at 25,0 °C)

Measuring point	Average value of reference humidity and temperature at the measuring point		Minimum value of reference humidity and temperature at the measuring point		Maximum value of reference humidity and temperature at the measuring point		Measurement uncertainty ²⁾	
	°C	% RH	°C	% RH	°C	% RH	°C	% RH
L1P1	25,06	82,76	24,94	82,32	25,50	83,72	0,60	2,31
L1P2	25,06	81,31	24,93	80,80	25,50	82,20	0,60	3,05

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

²⁾The measurement uncertainty does not include into account the spatial inhomogeneity of the chamber.

PARAMETERS CHARACTERIZING THE CHAMBER (at temperature setting : -18,0 °C)

Reference value	Inhomogeneity	Time stability ³⁾	Radiation effect
°C	°C	°C	°C
Measuring location L2, Point P1			
-18,55	0,28	1,10	-

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾Determined within 60 minutes

PARAMETERS CHARACTERIZING THE CHAMBER (at humidity setting: 90,0 % at 19,0 °C)

Reference value	Inhomogeneity	Time stability ³⁾	Radiation effect
°C	% RH	°C	% RH
Measuring location L2, Point P1			
19,20	92,51	0,01	1,34
		0,08	0,85
			0,02

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾Determined within 60 minutes

PARAMETERS CHARACTERIZING THE CHAMBER (at humidity setting: 90,0 % at 21,0 °C)

Reference value	Inhomogeneity	Time stability ³⁾	Radiation effect
°C	% RH	°C	% RH
Measuring location L2, Point P1			
21,23	90,52	0,01	1,38
		0,16	3,98
			0,02

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾Determined within 60 minutes

Authorized by:
M.Sc.Eng. Marta Piotrowska

Additional information:
Document signed with a qualified electronic signature.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 158

Date of issue: 12th June 2025

Certificate No: 182/6/25

Page 4/4

PARAMETERS CHARACTERIZING THE CHAMBER (at humidity setting: 80,0 % at 25,0 °C)


Reference value		Inhomogeneity		Time stability ³⁾		Radiation effect
°C	% RH	°C	% RH	°C	% RH	°C
Measuring location L2, Point P1						
25,06	81,31	0,00	1,45	0,44	0,89	0,02

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾Determined within 60 minutes

REVIEW DRAWING

Distribution of measurement points during calibration in the temperature setting: -18,0 °C and in the relative humidity setting: 90,0 %, at temperature 19,0 °C; 90,0 % at temperature 21,0 °C; 80,0 % at 25,0 °C; The drawing does not reflect the actual proportions.



Measuring point	Description of measurement points	Sensor No.
L1P1	The sensor is placed 1100 mm from the left and 2700 mm from the back wall, at a height of 1140 mm.	B/252/W/S/PW/P B/21/W/S/PW/P
L1P2	The sensor is placed 1140 mm from the left and 2700 mm from the back wall, at a height of 1140 mm - the reference sensor.	B/253/W/S/PW/P B/16/W/S/PW/P

Authorized by:
M.Sc.Eng. Marta Piotrowska

Additional information:
Document signed with a qualified electronic signature.


**TENSLAB Spółka z ograniczoną odpowiedzialnością
LABORATORIUM WZORCUJĄCE**

Siedziba: 30-554 GDAŃSK ul. Śnieżna 5
Laboratorium: 81-336 GDYNIA ul. Kadłubowców 4
Kontakt: www.tenslab.pl tel: +48 58 380 06 44



AP 180

Calibration laboratory accredited by Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA that include recognition of calibration certificates.
Accreditation No AP 180.



Baltic Fire Laboratory
BFL S/N: BFL120
Calib.Date: 13.06.2025
Next Calib.: 13.06.2027

CALIBRATION CERTIFICATE

Date of issue: 13 June 2025

Certificate No: 1061/2025/3

Page 1/ 2

**OBJECT OF
CALIBRATION**

Digital pressure gauge, Type: MD0906, Serial number: 2206201176750, Producer: Gesa, Resolution: 0,01 bar, Measuring range: 0,0+100,0 bar, Class: 0,1, Identification number: no data.

CUSTOMER

BALTIC FIRE LABORATORY SPÓŁKA Z O. O.
ul. Banińska 8
80-209 Tuchom

**PLACE OF
CALIBRATION**

TENSLAB Sp. z o.o. Laboratorium Wzorcujące
ul. Kadłubowców 4
81-336 Gdynia

**CALIBRATION
METHOD**

Procedura pomiarowa - Ciśnieniomierze elektroniczne - PP TENSLAB/CE/1/18 date of issue 20.07.2018

**ENVIRONMENTAL
CONDITIONS**

Ambient temperature: (21,8 ± 22,4) °C

**DATE OF
CALIBRATION**

13 June 2025

TRACEABILITY

This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).

**CALIBRATION
RESULTS**

The results have been presented on page(s) 2 of this certificate including uncertainty of measurement. The results relate only to the object of calibration.

**UNCERTAINTY OF
MEASUREMENT**

Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor $k = 2$.

This Calibration Certificate can be shown or copy only as a whole.
This document has been digitally signed.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 180

Date of issue: 13 June 2025

Certificate No: 1061/2025/3

Page 2/ 2

**CALIBRATION
RESULTS**

Calibration results are the following:

Determination of measurement error:

No	Measured pressure value	Reference pressure value	Measurement error	Hysteresis	Measurement uncertainty
bar					
1.	0,00	0,00	0,00	0,00	0,04
2.	5,00	5,03	-0,03	0,01	0,04
3.	8,00	8,03	-0,03	0,01	0,04
4.	16,00	16,03	-0,03	0,01	0,04
5.	30,00	30,02	-0,02	0,02	0,04
6.	40,00	40,02	-0,02	0,02	0,04
7.	50,00	50,02	-0,02	0,02	0,04
8.	60,00	60,02	-0,02	0,02	0,04
9.	75,00	75,01	-0,01	0,02	0,04
10.	90,00	90,01	-0,01	0,02	0,04
11.	100,00	100,02	-0,02	0,01	0,04
12.	-	-	-	-	-
13.	-	-	-	-	-
14.	-	-	-	-	-
15.	-	-	-	-	-

Medium: oil

Authorized by:

 Technical Manager
 Paula Keister-Lyko

END OF CALIBRATION CERTIFICATE


LABORATORIUM USŁUGOWO-BADAWCZE
„BIOCHEMIK” Sp. z o.o.
Pracownia Wzorcowania

ul. Strefowa 15, 64 – 920 Pila

Mob. +48 727 630 955, e-mail: wzorcowanie@biochemik.pl, www.wzorcowanie.biochemik.pl

 Calibration laboratory accredited by
 Polish Centre of Accreditation, a signatory to EA MLA and ILAC MRA
 that include recognition of calibration certificates.

Accreditation No AP 158



AP 158



CALIBRATION CERTIFICATE

 Date of issue: 28th November 2025

Certificate No: 848/11/25

Page 1/4

**OBJECT OF
CALIBRATION**

 Thermostatic chamber
 Manufacturer: WAMED
 Type: SUP-200G
 Serial number: -
 User marking: BFL 140
 Resolution: 0,1 °C
 Workspace dimensions: (width x depth x height): (550 x 400 x 1030) mm

 Baltic Fire Laboratory
 BFL S/N: BFL140
 Calib.Date: 17.11.2025
 Next Calib.: 17.11.2027

APPLICANT

 Baltic Fire Laboratory Sp. z o.o.
 ul. Banińska 8
 80-209 Tuchom, Poland

**PLACE OF
CALIBRATION**

 Baltic Fire Laboratory Sp. z o.o.
 ul. Banińska 8
 80-209 Tuchom, Poland

**CALIBRATION
METHOD**

Calibration has been carried out according to the calibration procedure „Climatic and thermostatic chambers”, PW-06, 7th edition of 02 January 2022.

**ENVIRONMENTAL
CONDITIONS**

 Temperature: (16,8 ÷ 18,3) °C
 Relative humidity: (41,9 ÷ 44,7) %

**DATE OF
CALIBRATION**

 17th November 2025

TRACEABILITY

This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).

**CALIBRATION
RESULTS**

The results have been presented on page 2/4 of this certificate including uncertainty of measurement. The results apply only to the calibration object.

**UNCERTAINTY OF
MEASUREMENT**

 Uncertainty of measurement has been evaluated in compliance EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95% and the coverage factor $k=2$.

 Head of the Calibration
 Laboratory



M.Sc.Eng. Paweł Janik

This certificate may be presented or copied as a whole document only.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 158

Date of issue: 28th November 2025

Certificate No: 848/11/25

Page 2/4

CALIBRATION RESULTS

Calibration results are the following:

Set value of the test object	Average measured reference value	Average temperature indication of the test object	Measurement error	Measurement uncertainty ¹⁾
t	t_w	t_x	Δt	U
°C	°C	°C	°C	°C
Measuring location L1, Point P2				
23,0	22,4	23,0	0,6	1,1
49,0	50,7	49,1	-1,6	0,9
104,0	105,6	104,0	-1,6	2,1
150,0	146,6	150,1	3,5	3,2

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

¹⁾The measurement uncertainty include into account the spatial inhomogeneity of the chamber.

TEMPERATURE DISTRIBUTION IN THE CHAMBER (at temperature setting: 23,0 °C)

Measuring point	Average value reference temperature at the measuring point	Minimum value of the reference temperature at the measuring point	Maximum value of the reference temperature at the measuring point	Measurement uncertainty ³⁾
	°C	°C	°C	°C
L1P1	22,35	21,85	22,98	1,01
L1P2	22,39	21,88	23,02	1,01

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾The measurement uncertainty does not include into account the spatial inhomogeneity of the chamber.

TEMPERATURE DISTRIBUTION IN THE CHAMBER (at temperature setting: 49,0 °C)

Measuring point	Average value reference temperature at the measuring point	Minimum value of the reference temperature at the measuring point	Maximum value of the reference temperature at the measuring point	Measurement uncertainty ³⁾
	°C	°C	°C	°C
L1P1	50,66	50,20	50,93	0,80
L1P2	50,73	50,20	51,02	0,86

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾The measurement uncertainty does not include into account the spatial inhomogeneity of the chamber.

TEMPERATURE DISTRIBUTION IN THE CHAMBER (at temperature setting: 104,0 °C)

Measuring point	Average value reference temperature at the measuring point	Minimum value of the reference temperature at the measuring point	Maximum value of the reference temperature at the measuring point	Measurement uncertainty ³⁾
	°C	°C	°C	°C
L1P1	105,23	104,44	106,42	2,01
L1P2	105,64	104,82	106,87	2,04

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾The measurement uncertainty does not include into account the spatial inhomogeneity of the chamber.

Authorized by:
M.Sc.Eng. Marta Piotrowska

Additional information:

Document signed with a qualified electronic signature.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 158

Date of issue: 28th November 2025

Certificate No: 848/11/25

Page 3/4

TEMPERATURE DISTRIBUTION IN THE CHAMBER (at temperature setting: 150,0 °C)

Measuring point	Average value of the reference temperature at the measuring point °C	Minimum value of the reference temperature at the measuring point °C	Maximum value of the reference temperature at the measuring point °C	Measurement uncertainty ³⁾ °C
L1P1	146,49	145,34	148,75	3,19
L1P2	146,58	145,52	148,82	3,17

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

²⁾The measurement uncertainty does not include into account the spatial inhomogeneity of the chamber.

PARAMETERS CHARACTERIZING THE CHAMBER (at temperature setting: 23,0 °C)

Reference value °C	Inhomogeneity °C	Time stability ³⁾ °C	Radiation effect °C
Measuring location L1, Point P2			
22,39	0,04	0,63	0,00

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾Determined within 60 minutes

PARAMETERS CHARACTERIZING THE CHAMBER (at temperature setting: 49,0 °C)

Reference value °C	Inhomogeneity °C	Time stability ³⁾ °C	Radiation effect °C
Measuring location L1, Point P2			
50,73	0,07	0,53	0,00

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾Determined within 60 minutes

PARAMETERS CHARACTERIZING THE CHAMBER (at temperature setting: 104,0 °C)

Reference value °C	Inhomogeneity °C	Time stability ³⁾ °C	Radiation effect °C
Measuring location L1, Point P2			
105,64	0,41	1,23	0,00

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾Determined within 60 minutes

PARAMETERS CHARACTERIZING THE CHAMBER (at temperature setting: 150,0 °C)

Reference value °C	Inhomogeneity °C	Time stability ³⁾ °C	Radiation effect °C
Measuring location L1, Point P2			
146,58	0,09	2,24	0,00

Calibration results refer to the International Temperature Scale of 1990 (ITS – 90)

³⁾Determined within 60 minutes

Authorized by:
M.Sc.Eng. Marta Piotrowska

Additional information:
Document signed with a qualified electronic signature.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 158


 Date of issue: 28th November 2025

Certificate No: 848/11/25

Page 4/4

REVIEW DRAWING

Distribution of measurement points during calibration in the temperature setting: 23,0 °C, 49,0 °C, 104,0 °C, 150,0 °C. The drawing does not reflect the actual proportions.



Measuring point	Description of measurement points	Sensor No.
L1P1	The sensor is placed 225 mm from the left and 200 mm from the back wall, at a height of 530 mm.	B/137/W/S/PW/P
L1P2	The sensor is placed 275 mm from the right and 200 mm from the back wall, at a height of 530 mm - the reference sensor.	B/139/W/S/PW/P

 Authorized by:
 M.Sc.Eng. Marta Piotrowska

 Additional information:
 Document signed with a qualified electronic signature.

RADWAG Wagi Elektroniczne Witold Lewandowski
26-600 Radom, Toruńska 5
METROLOGY, RESEARCH AND CERTIFICATION CENTRE - MEASURING LABORATORY
26-600 Radom, Starowiejska 17A
tel. /48/ 386 64 70; fax.: /48/ 385 00 11



Calibration laboratory accredited by
Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No AP 069.



AP 069



CALIBRATION CERTIFICATE

Dispatch date: 11 October 2025

Certificate No: 13817/4771/25/E

Page: 1 / 2

OBJECT OF CALIBRATION	Non-automatic electronic weighing instrument - single range	
	Manufacturer	RADWAG Wagi Elektroniczne
	Type / symbol	WLC 30/F1/R
	Serial No	856846
	Capacity Max	30 kg
	Scale interval d	0,5 g
Baltic Fire Laboratory BFL S/N: BFL141 Calib.Date: 11.10.2025 Next Calib.: 11.10.2027		
APPLICANT	RADWAG Wagi Elektroniczne ul. Toruńska 5, 26-600 Radom	
USER	Baltic Fire Laboratory Sp. z o.o. ul. Banińska 8, 80-209 Tuchom	
PLACE OF CALIBRATION	RADWAG Wagi Elektroniczne Centrum Metrologii, Badań i Certyfikacji - Laboratorium Pomiarowe ul. Starowiejska 17 A, 26-600 Radom	
CALIBRATION METHOD	Calibration Procedure PW 01 rev. XV of 2 January 2025 based on EURAMET cg-18 v. 4.0 (11/2015).	
ENVIRONMENTAL CONDITIONS	Air temperature:	(23,58 + 23,62) ± 0,20 °C
	Relative air humidity:	(49,0 + 49,3) ± 1,1 %
DATE OF CALIBRATION	11 October 2025	
TRACEABILITY	This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).	
CALIBRATION RESULTS	The results have been presented on page 2 of this certificate including uncertainty of measurement.	
UNCERTAINTY OF MEASUREMENT	Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95% and the coverage factor $k = 2$.	



KIEROWNIK
Laboratorium Pomiarowego
Tomasz Jedrzejewski

This certificate may be presented or copied as a whole document only.

CALIBRATION CERTIFICATE issued by ACCREDITED LABORATORY No AP 069

Dispatch date: 11 October 2025

Certificate No: 13817/4771/25/E

Page: 2 / 2

CALIBRATION RESULTS

Calibration results presented below refer only to the calibration object described on the first page of the certificate.


 Adjustment device: external
 Adjustment weight: external

MEASUREMENT ERROR


Load L g	Load TARE L_T g	Mass reference m_{ref} g	Indication of instrument I g	Measurement error $E(I)$ g	Uncertainty of measurement $U(E)$ g
50		50,00007	50,0	0,0	0,7
7000		7000,00463	7000,5	0,5	0,9
15000		15000,0042	15000,5	0,5	1,5
22000		22000,00823	22000,0	0,0	2,0
30000		30000,0078	30000,0	0,0	2,7

REPEATABILITY
ECCENTRICITY

Load L g	Standard deviation s g
30000	0,27
Standard deviation calculated for 5 series of measurement.	

Position of load	Load L_{ecc} g	Area no	Indication of instrument I_{ecc} g	Indication difference ΔI_{ecc} g
	10000	1	10000,5	$ \Delta I_{ecc} _{max} 1,5$
		2	10001,0	0,5
		3	9999,5	-1,0
		4	9999,0	-1,5
		5	10001,0	0,5

Authorized by: Tomasz Jędrzejewski



Baltic Fire Laboratory
BFL S/N: BFL1002
Calib.Date: 31.10.2024
Next Calib.: 31.10.2026



Institute of Power Engineering – National Research Institute

MEASURING APPARATUS LABORATORY

02-981 Warszawa, ul. Augustówka 36

tel. (22) 34 51 438

lap@ien.com.pl

www.ien.com.pl

Calibration laboratory accredited by
Polish Centre for Accreditation, a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No AP 013



AP 013

**CALIBRATION CERTIFICATE**


Date of issue: 31 October 2024

Certificate No: 4100/2024/E

Page 1/3

OBJECT OF CALIBRATION	Temperature indicator model NI 9213, manufacturer NATIONAL INSTRUMENTS, serial number 187F11C, 16-channel with a temperature measuring chain cooperating with K-type thermocouples.
APPLICANT	Baltic Fire Laboratory Sp. z o.o. 80 – 209 Tuchom, ul. Banińska 8 Order No: MAP.4032.1332.2024
PLACE OF CALIBRATION	Baltic Fire Laboratory Sp. z o.o. 80 – 209 Tuchom, ul. Banińska 8
CALIBRATION METHOD	Calibration was performed according to the Measurement Procedure P4/MAP „Calibration apparatus for measuring temperature, pressure, electrical quantities and humidity on objects”, edition XII released on 30 May 2023.
ENVIRONMENTAL CONDITIONS	Ambient temperature (19,6 + 20,2) °C Relative humidity (65,3 + 72,2) %
DATE OF CALIBRATION	22 October 2024
TRACEABILITY	This certificate is issued under the agreement EA MLA in the field of calibration and provides traceability of measurement results to the International System of Units (SI).
CALIBRATION RESULTS	The results have been presented on page 2/3 + 3/3 of this certificate including uncertainty of measurement. The result apply to the object of calibration only.
UNCERTAINTY OF MEASUREMENT	Uncertainty of measurement has been evaluated in compliance with EA-4/02 M:2022. The expanded uncertainty assigned corresponds to a coverage probability of 95 % and the coverage factor $k = 2$.



KIEROWNIK
Laboratorium Aparatury Pomiarowej

mgr Inż. Roman Witkowski

This certificate may be presented or copied as a whole document only.



AP 201



Guenther Polska Sp. z o.o.
ul. Wrocławska 27C 55-095 Długołęka
Laboratorium wzorcowujące akredytowane
przez Polskie Centrum Akredytacji,
sygnatariusza porozumień EA MLA i ILAC MRA
dotyczących wzajemnego uznawania świadectw
Nr akredytacji AP 201
Calibration laboratory accredited by
Polish Centre for Accreditation,
a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No. AP 201



ŚWIADECTWO WZORCOWANIA NR: PA2506021

CALIBRATION CERTIFICATE NO.

Data wydania: 24.06.2025

Release date

ZGŁASZAJĄCY
Applicant

BALTIC FIRE LABORATORY Sp. z o. o.
ul. Banińska 8
80-209 Tuchom

MIEJSCE WZORCOWANIA
Place of calibration

Guenther Polska Sp. z o.o.
ul. Wrocławska 27c
55-095 Długołęka

NR ZAMÓWIENIA ZEW.
External order No.

zam. z dnia 04.06.2025

NR ZAMÓWIENIA WEW.
Internal order No.

ZS 053/06/2025

OBIEKT WZORCOWANIA
Calibration object

Nazwa: Name	Czujnik termoelektryczny Thermocouple Non-expendable Ø : 1,5 mm	Typ: Type	1x NiCr-NiAl/K
----------------	--	--------------	----------------

Kod artykułu: Article code	20053110	Producent/Model: Manufacturer/Model	Guenther
-------------------------------	----------	--	----------

DATA WZORCOWANIA
Calibration date

20-23.06.2025

RODZAJ WZORCOWANIA
Calibration type

Rekalibracja / Recalibration

METODA WZORCOWANIA
Calibration method

Wzorcowanie zostało przeprowadzone zgodnie z procedurą QMV9.01.01 Wzorcowanie termoelementu metodą porównawczą ver. 3.0 z dnia 12.10.2021 (opracowanej na podstawie ASTM E220-19).

The calibration has been performed in accordance with procedure QMV9.01.01 Calibration of the thermocouple by the comparison method ver. 3.0 of 12.10.2021 (based on the ASTM E220-19).

SPÓJNOŚĆ POMIAROWA
Measurement traceability

Świadectwo potwierdza spójność wyników pomiarów z jednostkami miar Międzynarodowego Układu Jednostek Miar (SI).

The certificate provides traceability of measurement results to the International System of Units (SI).

NIEPEWNOŚĆ POMIARU
Measurement uncertainty

Niepewność pomiaru została określona zgodnie z dokumentem EA-4/02 M:2022. Podane wartości niepewności stanowią niepewności rozszerzone przy prawdopodobieństwie rozszerzenia ok. 95% i współczynnika rozszerzenia $k = 2$.

The measurement uncertainty has been determined in accordance with EA-4/02 M: 2022. The reported uncertainties of measurement are stated as the expanded uncertainties corresponded to a coverage probability of approximately 95% and a coverage factor $k = 2$.

UWAGI
Comments

Podane wyniki wzorcowania dotyczą wyłącznie wzorcowanego obiektu.
Calibration results apply only calibrated object.
W razie wątpliwości ważny jest oryginalny polski tekst.
If there are any doubts, the Polish version shall prevail.
Niniejsze świadectwo może być okazywane lub kopiowane tylko w całości.
This certificate may be presented or copied as a whole document only.

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 C Tel. +48 (0) 71 / 252 70 70
55-095 Długołęka, Polska Faks +48 (0) 71 / 252 70 71
biuro@guenther.com.pl www.guenther.com.pl

NIP 885-19-06-589
REGON 220702531
KRS 0000300821

BNP PARIBAS BANK POLSKA S.A.
PLN 95 1750 1044 0000 0000 2056 6722
EUR 61 1750 1044 0000 0000 2056 6757 Kod SWIFT/BIC: PFBAPL33XXX

Strona 1 / 4
Page


ŚWIADECTWO WZORCOWANIA NR: PA2506021

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

UŻYTE WZORCE I URZĄDZENIA POMIAROWE

Reference probes and measuring devices

WZORZEC

Reference probes

T ≤ 250 °C	Pt100 / 4	SW 44/35/INTIBS-LWT/2025	L-R-0025	07-04-2026
T > 250 °C	Pt100 / 4	4166/2024	L-R-0021	28-10-2025
T > 500 °C	PtRh10-PtS	PA2411069/1	L-T-0040	19-11-2025

URZĄDZENIE POMIAROWE

Measuring instrument

T ≤ 58 °C	Keysight	2672/2024	L-M-0007	14-07-2025
T > 58 °C	Keysight	76/2025	L-M-0008	16-01-2026
T > 250 °C	Keysight	2672/2024	L-M-0007	14-07-2025
T > 500 °C	Keysight	4990/2024	L-M-0009	22-12-2025

ELEMENT GRZEJNY

Temperature source

T ≤ 58 °C	Fluke	głębokość zanurzenia = 150 mm	L-H-0002
T > 58 °C	Fluke	głębokość zanurzenia = 150 mm	L-H-0003
T > 500 °C	Additel	głębokość zanurzenia = 150 mm	L-H-0011

Immersion depth

SPOINA ODNIESIENIA 0 °C

Cold junction

T ≤ 58 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0002
T > 58 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0004
T > 250 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0002

Immersion depth

WARUNKI ŚRODOWISKOWE

Environmental conditions

Temperatura otoczenia	(21,9 ± 24,7) °C
<i>Ambient temperature</i>	
Wilgotność względna	(29,9 ± 44,3) %
<i>Humidity</i>	

ŚWIADECTWO WZORCOWANIA NR: PA2506021

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

WYNIKI WZORCOWANIA
Results of calibration

CZUJNIK TYPU Sensor type	NOMINALNA Nominal	TEMPERATURA Temperatura ODNIESIENIA Reference	OBLICZONA Calculated	ZMIERZONE NAPIĘCIE TERMEOELEKTRYCZNE Thermovoltage	BŁĄD POMIARU Measurement error	NAJWIEKSZY BŁĄD DOPUSTYCALNY Maximum permissible error	NIEPENNOŚĆ POMIARU Measurement uncertainty	UWAGI Comments
	[°C]			[mV]	[°C]	± [°C]	± [°C]	
K	10,0	10,0	9,9	0,394	-0,1	1,5	0,7	1851220/1 PA2506021/1
	58,0	57,9	57,5	2,333	-0,4	1,5	0,7	
	250,0	250,1	249,1	10,118	-1,0	1,5	0,7	
	500,0	498,2	497,2	20,524	-1,0	2,0	0,7	
	950,0	950,5	948,3	39,247	-2,2	3,8	1,8	
K	10,0	10,0	9,8	0,391	-0,2	1,5	0,7	1851220/5 PA2506021/2
	58,0	57,9	57,6	2,338	-0,3	1,5	0,7	
	250,0	250,0	250,0	10,153	0,0	1,5	0,7	
	500,0	498,2	497,8	20,551	-0,4	2,0	0,7	
	950,0	950,0	949,5	39,295	-0,5	3,8	1,8	
K	10,0	10,0	9,8	0,391	-0,2	1,5	0,7	1851220/10 PA2506021/3
	58,0	57,9	57,6	2,337	-0,3	1,5	0,7	
	250,0	250,0	249,7	10,140	-0,3	1,5	0,7	
	500,0	498,2	497,4	20,532	-0,8	2,0	0,7	
	950,0	950,0	947,5	39,214	-2,5	3,8	1,8	
K	10,0	10,0	9,8	0,393	-0,2	1,5	0,7	1851220/15 PA2506021/4
	58,0	57,9	57,6	2,337	-0,3	1,5	0,7	
	250,0	250,0	250,4	10,171	0,4	1,5	0,7	
	500,0	499,5	499,7	20,632	0,2	2,0	0,7	
	950,0	950,0	950,3	39,325	0,3	3,8	1,8	
K	10,0	10,0	9,9	0,394	-0,1	1,5	0,7	1851220/22 PA2506021/5
	58,0	57,9	57,4	2,330	-0,5	1,5	0,7	
	250,0	250,0	249,2	10,121	-0,8	1,5	0,7	
	500,0	498,2	497,0	20,515	-1,2	2,0	0,7	
	950,0	950,7	949,4	39,291	-1,3	3,8	1,8	
K	10,0	10,0	9,9	0,393	-0,1	1,5	0,7	1851220/31 PA2506021/6
	58,0	57,9	57,5	2,333	-0,4	1,5	0,7	
	250,0	250,0	249,6	10,138	-0,4	1,5	0,7	
	500,0	498,2	497,5	20,536	-0,7	2,0	0,7	
	950,0	950,0	949,4	39,289	-0,6	3,8	1,8	
K	10,0	10,0	9,9	0,395	-0,1	1,5	0,7	1851220/33 PA2506021/7
	58,0	57,9	57,6	2,337	-0,3	1,5	0,7	
	250,0	250,0	249,2	10,123	-0,8	1,5	0,7	
	500,0	499,5	499,1	20,605	-0,4	2,0	0,7	
	950,0	950,0	950,1	39,317	0,1	3,8	1,8	

Data wzorcowania: 20-23.06.2025

Calibration date

Strona 3 / 4

Page

GUENTHER Polska Sp. z o.o.

ul. Wrocławia 27 C Tel. +48 (0)71 / 352 70 70
55-055 Długosza Faks +48 (0)71 / 352 70 71
Polska biuro@guenther.com.pl - www.guenther.com.pl

NP 895-19-06-589 BNP PARIBAS BANK POLSKA S.A.
REGON 143702531 PLN 86 1793 1044 0000 0000 2050 6722
KRS 000090662 EUR 61 1750 1084 0000 0000 2050 6757 Kod SWIFT/BIC: PBPAPL3XXX


ŚWIADECTWO WZORCOWANIA NR: PA2506021

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

CZUJNIK TYP Sensor type	TEMPERATURA Temperatura ODNIESIENIA Reference		OBLICZONA Calculated	ZMIERZONE NAPIĘCIE TERMIELEKTRYCZNE Thermovoltage	BŁĄD POMIARU Measurement error	NAJWIĘKSZY BŁĄD DOPUSZCZALNY Maximum permissible error	NIEPEWNOŚĆ POMIARU Measurement uncertainty	UWAGI Comments
	NOMINALNA Nominal	[°C]						
K	10,0	10,0	9,9	0,394	-0,1	1,5	0,7	185122038 PA2506021/8
	58,0	57,9	57,4	2,329	-0,5	1,5	0,7	
	250,0	250,1	248,8	10,106	-1,3	1,5	0,7	
	500,0	498,2	496,8	20,507	-1,4	2,0	0,7	
	950,0	950,5	948,3	39,245	-2,2	3,8	1,8	
K	10,0	10,0	9,8	0,391	-0,2	1,5	0,7	185122039 PA2506021/9
	58,0	57,9	57,6	2,338	-0,3	1,5	0,7	
	250,0	250,0	250,2	10,162	0,2	1,5	0,7	
	500,0	499,5	499,6	20,626	0,1	2,0	0,7	
	950,0	950,0	949,9	39,309	-0,1	3,8	1,8	
K	10,0	10,0	9,9	0,395	-0,1	1,5	0,7	185122048 PA2506021/10
	58,0	57,9	57,6	2,334	-0,3	1,5	0,7	
	250,0	250,1	249,2	10,121	-0,9	1,5	0,7	
	500,0	498,2	496,8	20,507	-1,4	2,0	0,7	
	950,0	950,5	948,7	39,264	-1,8	3,8	1,8	

UWAGI

Comments

- Temperatura nominalna: punkt nastawy wynikający z zamówienia.
Nominal temperature: set point specified in the order.
- Temperatura odniesienia: średnia wartość z pomiarów odczytanych z wzorca.
Reference temperature: average value of the measurement results read from the reference standard.
- Temperatura obliczona: wartość temperatury obliczona na podstawie normy ASTM E230M-23a z wartości zmierzonej napięcia termoelektrycznego obiektu.
Calculated temperature: temperature value calculated from the measured thermovoltage of the calibrated object on the basis of the ASTM E230M-23a.
- Błąd pomiaru = Temperatura obliczona - Temperatura odniesienia
Measurement error = Calculated temperature - Reference temperature
- Największy błąd dopuszczalny: dopuszczalny błąd pomiaru temperatury według klasy 1 wg PN EN 60584-1:2014 pkt 5.1.
Maximum permissible error: according to the Class 1 acc. to PN EN 60584-1: 2014 point 5.1.
- Stwierdzenie zgodności z daną specyfikacją podjęto na podstawie zasady prostej akceptacji ILAC G8:09/2019 w=0 TUR>3:1.
Ryzyko specyficzne błędnej akceptacji/ błędnego odrzucenia wynosi 50%.
- Compliance with the principles of the simple acceptance according to the ILAC G8:09/2019 w=0 TUR >3:1. The specific risk of a false acceptance/false rejection can be up to 50%.

ZGODNOŚĆ Z WYMAGANIAMI

Conformity with requirements

Na podstawie wszystkich wyników wzorcowania stwierdzono zgodność z wymaganiami PN-EN 60584-1:2014-04 pkt 5.1 dla klasy 1. Conformance with PN-EN 60584-1:2014-04 pt 5.1 class 1 was found on the basis of all calibration results.

Wykonał: A. Grykałowska

Made by

Inżynier Laboratorium

Laboratory Engineer

dr inż. Agnieszka Grykałowska

Autoryzował:

Approved

Kierownik Laboratorium

Head of Laboratory

Podpis:

Signature

mgr inż. Elwira Stojka

KONIEC/ End

Data wzorcowania: 20-23.06.2025

Calibration date

Strona 4 / 4

Page



AP 201



Guenther Polska Sp. z o.o.
ul. Wrocławska 27C 55-095 Długołęka
Laboratorium wzorcowujące akredytowane
przez Polskie Centrum Akredytacji,
sygnatariusza porozumień EA MLA i ILAC MRA
dotyczących wzajemnego uznawania świadectw
Nr akredytacji AP 201
Calibration laboratory accredited by
Polish Centre for Accreditation,
a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No. AP 201


GUENTHER Polska
member of the GUENTHER Group

Technologia pomiaru temperatury
Niezawodna • Dokładna • Certyfikowana

ŚWIADECTWO WZORCOWANIA NR: PA2507022

CALIBRATION CERTIFICATE NO.

Data wydania: 14.07.2025

Release date

ZGŁASZAJĄCY
Applicant

BALTIC FIRE LABORATORY Sp. z o. o.
ul. Banińska 8
80-209 Tuchom

MIEJSCE WZORCOWANIA
Place of calibration

Guenther Polska Sp. z o.o.
ul. Wrocławska 27c
55-095 Długołęka

NR ZAMÓWIENIA ZEW.
External order No.

zam. z dnia 02.07.25

NR ZAMÓWIENIA WEW.
Internal order No.

ZS 023/07/2025
OBIEKT WZORCOWANIA
Calibration object

Nazwa: Name	Czujnik termoelektryczny Thermocouple Non-expendable Ø : 1,5 mm	Typ: Type	1x NiCr-NiAl/K
Kod artykułu: Article code	20053110	Nr fabryczny: Serial No.	1851220/13 1851220/17 1851220/21 1851220/24 1851220/36 1851220/37 1851220/42
		Producent/Model: Manufacturer/Model	Guenther

DATA WZORCOWANIA
Calibration date

11-14.07.2025

RODZAJ WZORCOWANIA
Calibration type

Rekalibracja / Recalibration

METODA WZORCOWANIA
Calibration method

Wzorcowanie zostało przeprowadzone zgodnie z procedurą QMV9.01.01 Wzorcowanie termoelementu metodą porównawczą ver. 3.0 z dnia 12.10.2021 (opracowanej na podstawie ASTM E220-19).

The calibration has been performed in accordance with procedure QMV9.01.01 Calibration of the thermocouple by the comparison method ver.3.0 of 12.10.2021 (based on the ASTM E220-19).

SPÓJNOŚĆ POMIAROWA
Measurement traceability

Świadectwo potwierdza spójność wyników pomiarów z jednostkami miar Międzynarodowego Układu Jednostek Miar (SI).

The certificate provides traceability of measurement results to the International System of Units (SI).

NIEPEWNOŚĆ POMIARU
Measurement uncertainty

Niepewność pomiaru została określona zgodnie z dokumentem EA-4/02 M:2022. Podane wartości niepewności stanowią niepewności rozszerzone przy prawdopodobieństwie rozszerzenia ok. 95% i współczynniku rozszerzenia $k = 2$.

The measurement uncertainty has been determined in accordance with EA-4/02 M: 2022. The reported uncertainties of measurement are stated as the expanded uncertainties corresponded to a coverage probability of approximately 95% and a coverage factor $k = 2$.

UWAGI
Comments

Podane wyniki wzorcowania dotyczą wyłącznie wzorcowanego obiektu.
Calibration results apply only calibrated object.
W razie wątpliwości ważny jest oryginalny polski tekst.
If there are any doubts, the Polish version shall prevail.
Niniejsze świadectwo może być okazywane lub kopiowane tylko w całości.
This certificate may be presented or copied as a whole document only.

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 C Tel. +48 (0)71 / 352 70 70
55-095 Długołęka Polska Faks +48 (0)71 / 352 70 71
biuro@guenther.com.pl · www.guenther.com.pl

NIP 695-19-05-589 BNP PARIBAS BANK POLSKA S.A.
REGON 140702531 PLN 36 1790 1064 0000 0000 2055 6722
KRS 0000300621 BUR 61 1750 1064 0006 0000 2055 5757 Kod SWIFT/BIC: PPABLPXXXX

Strona 1 / 4
Page


ŚWIADECTWO WZORCOWANIA NR: PA2507022

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

UŻYTE WZORCE I URZĄDZENIA POMIAROWE

Reference probes and measuring devices

WZORZEC

Reference probes

T _≤ 58 °C	Pt100 / 4	SW 43/35/INTIBS-LWT/2025	L-R-0024	07-04-2026
T _{>} 58 °C	Pt100 / 4	SW 44/35/INTIBS-LWT/2025	L-R-0025	07-04-2026
T _{>} 250 °C	Pt100 / 4	4166/2024	L-R-0021	28-10-2025
T _{>} 500 °C	PIRh10-PI/S	1012-KL-35028-24	L-T-0039	30-10-2025

URZĄDZENIE POMIAROWE

Measuring instrument

T _≤ 58 °C	Keysight	Z4.Z42.4180.124.2024.3381.1	L-M-0013	20-10-2025
T _{>} 58 °C	Keysight	Z4.Z42.4180.49.2025.1433.1	L-M-0011	27-04-2026
T _{>} 250 °C	Keysight	Z4.Z42.4180.124.2024.3381.1	L-M-0013	20-10-2025
T _{>} 500 °C	Keysight	76/2025	L-M-0008	16-01-2026

ELEMENT GRZEJNY

Temperature source

T _≤ 58 °C	Fluke	głębokość zanurzenia = 150 mm	L-H-0002
T _{>} 58 °C	Fluke	głębokość zanurzenia = 200 mm	L-H-0006
T _{>} 500 °C	Additel	głębokość zanurzenia = 150 mm	L-H-0011

Immersion depth

SPOINA ODNIESIENIA 0 °C

Cold junction

T _≤ 500 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0003
T _{>} 500 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0004

Immersion depth

WARUNKI ŚRODOWISKOWE

Environmental conditions

Temperatura otoczenia	(21,9 + 25,3) °C
Ambient temperature	
Wilgotność względna	(35,6 + 53,5) %
Humidity	

ŚWIADECTWO WZORCOWANIA NR: PA2507022

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201


Technologie pomiaru temperatury
Niezawodne . Dokładne . Certyfikowane

WYNIKI WZORCOWANIA

Calibration results

CEWNIK TYPU Sensor type	NOMINALNA Nominal	TEMPERATURA Temperatura ODNIESIENIA Reference	OBLICZONA Calculated	ZMIENIONE NAPIĘCIE TERMIELEKTRYCZNE Thermoelectric voltage	BLĄD POMIARU Measurement error	NAJWIĘKSZY BLĄD DOPUSZCZALNY Maximum permissible error	WIEKOWOŚĆ POMIARU Measurement uncertainty	UWAGI Comments
		[°C]		[mV]	[°C]	± [°C]	± [°C]	
K	10,0	10,1	10,1	0,403	0,0	1,5	0,7	PA2507022/1 1851220/13
	58,0	57,3	57,0	2,309	-0,3	1,5	0,7	
	250,0	249,5	249,5	10,132	0,0	1,5	0,7	
	500,0	499,6	499,9	20,642	0,3	2,0	0,7	
	950,0	950,4	947,8	39,228	-2,6	3,8	1,8	
K	10,0	10,0	9,9	0,396	-0,1	1,5	0,7	PA2507022/2 1851220/17
	58,0	57,7	57,2	2,320	-0,5	1,5	0,7	
	250,0	249,5	249,5	10,134	0,0	1,5	0,7	
	500,0	499,6	498,1	20,562	-1,5	2,0	0,7	
	950,0	950,7	949,6	39,297	-1,1	3,8	1,8	
K	10,0	10,0	9,9	0,395	-0,1	1,5	0,7	PA2507022/3 1851220/21
	58,0	57,7	57,2	2,317	-0,5	1,5	0,7	
	250,0	249,5	249,5	10,135	0,0	1,5	0,7	
	500,0	499,6	498,5	20,580	-1,1	2,0	0,7	
	950,0	950,8	949,8	39,305	-1,0	3,8	1,8	
K	10,0	10,0	9,9	0,394	-0,1	1,5	0,7	PA2507022/4 1851220/24
	58,0	57,7	57,4	2,329	-0,3	1,5	0,7	
	250,0	249,5	248,6	10,096	-0,9	1,5	0,7	
	500,0	499,6	500,2	20,655	0,6	2,0	0,7	
	950,0	950,7	950,2	39,322	-0,5	3,8	1,8	
K	10,0	10,0	10,0	0,397	0,0	1,5	0,7	PA2507022/5 1851220/36
	58,0	57,7	57,2	2,319	-0,5	1,5	0,7	
	250,0	249,5	248,6	10,099	-0,9	1,5	0,7	
	500,0	499,6	500,0	20,643	0,4	2,0	0,7	
	950,0	950,7	949,9	39,309	-0,8	3,8	1,8	
K	10,0	10,0	9,9	0,395	-0,1	1,5	0,7	PA2507022/6 1851220/37
	58,0	57,7	57,1	2,317	-0,6	1,5	0,7	
	250,0	249,5	249,3	10,125	-0,2	1,5	0,7	
	500,0	499,6	499,9	20,639	0,3	2,0	0,7	
	950,0	950,4	949,1	39,278	-1,3	3,8	1,8	
K	10,0	10,0	10,0	0,398	0,0	1,5	0,7	PA2507022/7 1851220/42
	58,0	57,7	57,1	2,316	-0,6	1,5	0,7	
	250,0	249,5	249,3	10,125	-0,2	1,5	0,7	
	500,0	499,6	499,5	20,623	-0,1	2,0	0,7	
	950,0	950,7	947,7	39,224	-3,0	3,8	1,8	

Data wzorcowania: 11-14.07.2025

Calibration date

Strona 3 / 4

Page

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 C
55-095 Długosieja
Polska
Tel. +48 (0) 71 / 352 70 70
Fax +48 (0) 71 / 352 70 71
biuro@guenther.com.pl · www.guenther.com.pl

NIP 695-19-06-589
REGON 020702531
KRS 0000305821

BNP PARIBAS BANK POLSKA S.A.
PLN 36 1750 1054 0000 0000 2055 6722
EUR 81 1750 1054 0000 0000 2055 6757 Kod SWIFT/BIC: FPABPLXXXX


ŚWIADECTWO WZORCOWANIA NR: PA2507022

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

UWAGI*Comments*

- Temperatura nominalna: punkt nastawy wynikający z zamówienia.
Nominal temperature: set point specified in the order.
- Temperatura odniesienia: średnia wartość z pomiarów odczytanych z wzorca.
Reference temperature: average value of the measurement results read from the reference standard.
- Temperatura obliczona: wartość temperatury obliczona na podstawie normy ASTM E230M-23a z wartości zmierzonej napięcia termoelektrycznego obiektu.
Calculated temperature: temperature value calculated from the measured thermovoltage of the calibrated object on the basis of the ASTM E230M-23a.
- Błąd pomiaru = Temperatura obliczona - Temperatura odniesienia
Measurement error = Calculated temperature - Reference temperature
- Największy błąd dopuszczalny: dopuszczalny błąd pomiaru temperatury według klasy 1 wg PN EN 60584-1:2014 pkt 5.1.
Maximum permissible error: according to the Class 1 acc. to PN EN 60584-1: 2014 point 5.1.
- Stwierdzenie zgodności z daną specyfikacją podjęto na podstawie zasady prostej akceptacji ILAC G8:09/2019 w=0 TUR>3:1.
Ryzyko specyficzne błędnej akceptacji/ błędnego odrzucenia wynosi 50%.
- Compliance with the principles of the simple acceptance according to the ILAC G8:09/2019 w=0 TUR >3:1. The specific risk of a false acceptance/false rejection can be up to 50%.

ZGODNOŚĆ Z WYMAGANIAMI*Conformity with requirements*

Na podstawie wszystkich wyników wzorcowania stwierdzono zgodność z wymaganiami PN-EN 60584-1:2014-04 pkt 5.1 dla klasy 1. Conformance with PN-EN 60584-1:2014-04 pt 5.1 class 1 was found on the basis of all calibration results.

Wykonał:
*Made by*Inżynier Laboratorium
*Laboratory Engineer*Podpis:
Signature

dr inż. Agnieszka Grykałowska

Autoryzował:
*Approved*7-ca Kierownika Laboratorium
*Deputy Head of Laboratory*Podpis:
Signature

mgr Joanna Zając

KONIEC/ End



AP 201



Guenther Polska Sp. z o.o.
ul. Wrocławska 27C 55-095 Długołęka
Laboratorium wzorcowujące akredytowane
przez Polskie Centrum Akredytacji,
sygnatariusza porozumienia EA MLA i ILAC MRA
dotyczących wzajemnego uznawania świadectw
Nr akredytacji AP 201
Calibration laboratory accredited by
Polish Centre for Accreditation,
a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No. AP 201


GUENTHER Polska
member of the GUENTHER Group

Technologie pomiaru temperatury
Niezawodna · Dokładna · Certyfikowana

ŚWIADECTWO WZORCOWANIA NR: PA2507023

CALIBRATION CERTIFICATE NO.

Data wydania: 14.07.2025

Release date

ZGŁASZAJĄCY

Applicant

BALTIC FIRE LABORATORY Sp. z o.o.

ul. Banińska 8
80-209 Tuchom

MIEJSCE
WZORCOWANIA

Place of calibration

Guenther Polska Sp. z o.o.

ul. Wrocławska 27c
55-095 Długołęka

NR ZAMÓWIENIA ZEW.

External order No.

zam. z dnia 02.07.25

NR ZAMÓWIENIA WEW.

Internal order No.

ZS 023/07/2025
OBIEKT
WZORCOWANIA

Calibration object

Nazwa:

Name

Czujnik termoelektryczny

Thermocouple

Non-expendable

Ø : 1,5 mm

Typ:

Type

1x NiCr-NiAl/K

Kod artykułu:

Article code

20053099

Nr fabryczny:

Serial No.

1821220/17
1821220/20
1821220/37

Producent/Model:

Manufacturer/Model

Guenther
DATA WZORCOWANIA

Calibration date

11-14.07.2025

RODZAJ
WZORCOWANIA

Calibration type

Rekalibracja / Recalibration

METODA
WZORCOWANIA

Calibration method

Wzorcowanie zostało przeprowadzone zgodnie z procedurą QMV9.01.01 Wzorcowanie termoelementu metodą porównawczą ver. 3.0 z dnia 12.10.2021 (opracowanej na podstawie ASTM E220-19).

The calibration has been performed in accordance with procedure QMV9.01.01 Calibration of the thermocouple by the comparison method ver. 3.0 of 12.10.2021 (based on the ASTM E220-19).

SPÓJNOŚĆ
POMIAROWA

Measurement traceability

Świadectwo potwierdza spójność wyników pomiarów z jednostkami miar Międzynarodowego Układu Jednostek Miar (SI).

The certificate provides traceability of measurement results to the International System of Units (SI).

NIEPEWNOŚĆ
POMIARU

Measurement uncertainty

Niepewność pomiaru została określona zgodnie z dokumentem EA-4/02 M:2022.

Podane wartości niepewności stanowią niepewności rozszerzone przy prawdopodobieństwie rozszerzenia ok. 95% i współczynniku rozszerzenia $k = 2$.

The measurement uncertainty has been determined in accordance with EA-4/02 M: 2022.

The reported uncertainties of measurement are stated as the expanded uncertainties corresponded to a coverage probability of approximately 95% and a coverage factor $k = 2$.

UWAGI

Comments

Podane wyniki wzorcowania dotyczą wyłącznie wzorcowanego obiektu.

Calibration results apply only calibrated object.

W razie wątpliwości ważny jest oryginalny polski tekst.

If there are any doubts, the Polish version shall prevail.

Niniejsze świadectwo może być okazywane lub kopiowane tylko w całości.

This certificate may be presented or copied as a whole document only.

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 C
55-095 Długołęka
Polska

Tel. +48 (0)71 / 352 70 70
Faks +48 (0)71 / 352 70 71
baltic@guenther.com.pl - www.guenther.com.pl

NIP 685-19-06-589
REGON 140702531
KRS 0000300821

BNP PARIBAS BANK POLSKA S.A.
PLN 36 1750 1064 0000 0000 2055 6722
EUR 61 1750 1064 0000 0000 2055 6737 Kod SWIFT/BIC: PABLPK3XXX

Strona 1 / 3

Page


DNV
Euronormowany
Białystok 1-10

ŚWIADECTWO WZORCOWANIA NR: PA2507023

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

UŻYTE WZORCE I URZĄDZENIA POMIAROWE

Reference probes and measuring devices

WZORZEC

Reference probes

T ≤ 58 °C	Pt100 / 4	SW 43/35/INTIBS-LWT/2025	L-R-0024	07-04-2026
T > 58 °C	Pt100 / 4	SW 44/35/INTIBS-LWT/2025	L-R-0025	07-04-2026
T > 250 °C	Pt100 / 4	4166/2024	L-R-0021	28-10-2025
T > 500 °C	PtRh10-Pt/S	1012-KL-35026-24	L-T-0039	30-10-2025

URZĄDZENIE POMIAROWE

Measuring instrument

T ≤ 58 °C	Keysight	Z4.Z42.4180.124.2024.3381.1	L-M-0013	20-10-2025
T > 58 °C	Keysight	4990/2024	L-M-0009	22-12-2025
T > 250 °C	Keysight	76/2025	L-M-0008	16-01-2026
T > 500 °C	Keysight	76/2025	L-M-0008	16-01-2026

ELEMENT GRZEJNY

Temperature source

T ≤ 58 °C	Fluke	głębokość zanurzenia = 150 mm	L-H-0002
T > 58 °C	Fluke	głębokość zanurzenia = 150 mm	L-H-0003
T > 250 °C	Fluke	głębokość zanurzenia = 200 mm	L-H-0006
T > 500 °C	Additel	głębokość zanurzenia = 150 mm	L-H-0011

Immersion depth

SPOINA ODNIESIENIA 0 °C

Cold junction

T ≤ 58 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0003
T > 58 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0002
T > 250 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0003
T > 500 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0002

Immersion depth

WARUNKI ŚRODOWISKOWE

Environmental conditions

Temperatura otoczenia (21,9 ± 25,3) °C

Ambient temperature

Wilgotność względna (35,6 ± 53,5) %

Humidity

ŚWIADECTWO WZORCOWANIA NR: PA2507023

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

Technologie pomiaru temperatury
Niezawodne · Dokładne · Certyfikowane

WYNIKI WZORCOWANIA

Calibration results

CZUJNIK TYPE Sensor type	NOMINALNA Nominal	TEMPERATURA Temperatura ODNIESIENIA Reference	OBLICZONA Calculated	ZMIERZONE NAPIĘCIE TERMIELEKTRYCZNE Thermovoltage	BŁĄD POMIARU Measurement error	NAJWIĘKSZY BŁĄD DOPUSZCZALNY Maximum permissible error	NIEPENNOŚĆ POMIARU Measurement uncertainty	UWAGI Comments
		[°C]		[mV]	[°C]	± [°C]	± [°C]	
K	10,0	10,0	10,0	0,398	0,0	1,5	0,7	PA2507023/1 1821220/17
	58,0	57,7	57,4	2,329	-0,3	1,5	0,7	
	250,0	250,7	250,1	10,157	-0,6	1,5	0,7	
	500,0	499,6	499,6	20,627	0,0	2,0	0,7	
	950,0	950,8	949,9	39,312	-0,9	3,8	1,8	
K	10,0	10,0	10,0	0,397	0,0	1,5	0,7	PA2507023/2 1821220/20
	58,0	57,7	57,4	2,330	-0,3	1,5	0,7	
	250,0	250,7	250,3	10,166	-0,4	1,5	0,7	
	500,0	499,6	499,2	20,612	-0,4	2,0	0,7	
	950,0	950,8	949,0	39,276	-1,8	3,8	1,8	
K	10,0	10,0	10,1	0,402	0,1	1,5	0,7	PA2507023/3 1821220/37
	58,0	57,7	57,3	2,322	-0,4	1,5	0,7	
	250,0	249,5	248,4	10,088	-1,1	1,5	0,7	
	500,0	499,6	497,6	20,541	-2,0	2,0	0,7	
	950,0	950,8	947,7	39,224	-3,1	3,8	1,8	

UWAGI

Comments

- Temperatura nominalna: punkt nastawy wynikający z zamówienia.
Nominal temperature: set point specified in the order.
- Temperatura odniesienia: średnia wartość z pomiarów odczytanych z wzorca.
Reference temperature: average value of the measurement results read from the reference standard.
- Temperatura obliczona: wartość temperatury obliczona na podstawie normy ASTM E230M-23a z wartości zmierzonej napięcia termoelektrycznego obiektu.
Calculated temperature: temperature value calculated from the measured thermovoltage of the calibrated object on the basis of the ASTM E230M-23a.
- Błąd pomiaru = Temperatura obliczona - Temperatura odniesienia
Measurement error = Calculated temperature - Reference temperature
- Największy błąd dopuszczalny: dopuszczalny błąd pomiaru temperatury według klasy 1 wg PN EN 60584-1:2014 pkt 5.1.
Maximum permissible error: according to the Class 1 acc. to PN EN 60584-1: 2014 point 5.1.
- Stwierdzenie zgodności z daną specyfikacją podjęto na podstawie zasady prostej akceptacji ILAC G8:09/2019 w=0 TUR>3:1.
- Ryzyko specyficzne błędnej akceptacji/ błędnego odrzucenia wynosi 50%.
- Compliance with the principles of the simple acceptance according to the ILAC G8:09/2019 w=0 TUR >3:1. The specific risk of a false acceptance/false rejection can be up to 50%.

ZGODNOŚĆ Z WYMAGANIAMI

Conformity with requirements

Na podstawie wszystkich wyników wzorcowania stwierdzono zgodność z wymaganiami PN-EN 60584-1:2014-04 pkt 5.1 dla klasy 1. Conformance with PN-EN 60584-1:2014-04 pt 5.1 class 1 was found on the basis of all calibration results.

Wykonał:

Made by

Inżynier Laboratorium
Laboratory Engineer

Autoryzował:

Approved

Z-ca Kierownika Laboratorium
Deputy Head of Laboratory

Podpis:

Signature

dr inż. Agnieszka Grykałowska

Podpis:

Signature

mgr Joanna Zajac

KONIEC/ End

Data wzorcowania: 11-14.07.2025

Calibration date

Strona 3 / 3

Page

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 D
55-095 Długołęka
Polska

Tel.: +48 (0)71 / 352 70 70
Faks: +48 (0)71 / 352 70 71

biuro@guenther.com.pl · www.guenther.com.pl

NIP 885-19-05-389
REGON 020702531
KRS 0000350821

BNP PARIBAS BANK POLSKA S.A.
PLN 36 1750 1054 0000 0009 2055 6722
EUR 61 1750 1054 0000 0000 2055 6757 Kod SWIFT/BIC: FBABPL33XXX




Guenther Polska Sp. z o.o.
ul. Wrocławska 27C 55-095 Długoleka
Laboratorium wzorcowujące akredytowane
przez Polskie Centrum Akredytacji,
sygnatariusza porozumień EA MLA i ILAC MRA
dotyczących wzajemnego uznawania świadectw
Nr akredytacji AP 201
Calibration laboratory accredited by
Polish Centre for Accreditation,
a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No. AP 201



ŚWIADECTWO WZORCOWANIA NR: PA2504047/2

CALIBRATION CERTIFICATE NO.

Data wydania: 14.04.2025

Release date

ZGŁASZAJĄCY
Applicant

BALTIC FIRE LABORATORY Sp. z o.o.
ul. Banińska 8
80-209 Tuchom

MIEJSCE
WZORCOWANIA
Place of calibration

Guenther Polska Sp. z o.o.
ul. Wrocławska 27C
55-095 Długoleka

NR ZAMÓWIENIA ZEW.
External order No.

Zam. z dnia 04.04.2025

NR ZAMÓWIENIA WEW.
Internal order No.

ZS 080/04/2025

OBIEKT
WZORCOWANIA
Calibration object

Nazwa: Czujnik termoelektryczny Typ: 1 x NiCr-NiAl/K
Name: Thermocouple
Non-expendable
Ø : 1,5 mm

Kod artykułu: 20053105 Nr fabryczny: 1831220/44
Article code: Serial No.

Szczegół: 126896 Producent/Model: Guenther
Batch No. Manufacturer/Model

DATA WZORCOWANIA
Calibration date

09.04.2025

RODZAJ
WZORCOWANIA
Calibration type

Rekalibracja / Recalibration

METODA
WZORCOWANIA
Calibration method

Wzorcowanie zostało przeprowadzone zgodnie z procedurą QMV9.01.01 Wzorcowanie termoelementu metodą porównawczą ver. 3.0 z dnia 12.10.2021 (opracowanej na podstawie ASTM E220-19).

The calibration has been performed in accordance with procedure QMV9.01.01 Calibration of the thermocouple by the comparison method ver.3.0 of 12.10.2021 (based on the ASTM E220-19).

SPÓJNOŚĆ
POMIAROWA
Measurement traceability

Świadectwo potwierdza spójność wyników pomiarów z jednostkami miar Międzynarodowego Układu Jednostek Miar (SI).

The certificate provides traceability of measurement results to the International System of Units (SI).

NIEPEWNOŚĆ
POMIARU
Measurement uncertainty

Niepewność pomiaru została określona zgodnie z dokumentem EA-4/02 M:2022. Podane wartości niepewności stanowią niepewności rozszerzone przy prawdopodobieństwie rozszerzenia ok. 95% i współczynniku rozszerzenia $k = 2$.

The reported uncertainty has been determined in accordance with EA-4/02 M: 2022. The reported uncertainties of measurement are stated as the expanded uncertainties corresponded to a coverage probability of approximately 95% and a coverage factor $k = 2$.

UWAGI
Comments

Podane wyniki wzorcowania dotyczą wyłącznie wzorcowanego obiektu.

Calibration results apply only calibrated object.

W razie wątpliwości ważny jest oryginalny polski tekst.

If there are any doubts, the Polish version shall prevail.

Niniejsze świadectwo może być okazywane lub kopiowane tylko w całości.

This certificate may be presented or copied as a whole document only.

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 C Tel. +48 (0)71 / 352 70 70
55-095 Długoleka Peka +48 (0)71 / 352 70 71
Polska biuro@guenther.com.pl - www.guenther.com.pl

NIP 895-19-08-589 BNP PARIBAS BANK POLSKA S.A.
REGON 00702531 PLN 36 1750 1064 0000 0000 2055 6732
KRS 0006390621 EUR 61 1750 1064 0000 0000 2055 6737 Kod SWIFT/BIC: PABSPKXXXX

Strona 1 / 3
Page


ŚWIADECTWO WZORCOWANIA NR: PA2504047/2

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

UŻYTE WZORCE I URZĄDZENIA POMIAROWE*Reference probes and measuring devices***WZORZEC***Reference probes*

T ≤ 58 °C	Pt100 / 4	Z2-Z21.4180.74.2023.2539.1	L-R-0004	06-09-2025
T > 58 °C	PtRh10-Pt/S	PA2411069/1	L-T-0040	19-11-2025

URZĄDZENIE POMIAROWE*Measuring instrument*

Keysight	Z4.Z42.4180.124.2024.3381.1	L-M-0013	20-10-2025
----------	-----------------------------	----------	------------

ELEMENT GRZEJNY*Temperature source*

T ≤ 58 °C	Fluke	głębokość zanurzenia = 150 mm	L-H-0002
T > 58 °C	Additel	głębokość zanurzenia = 150 mm	L-H-0011

*Immersion depth***SPOINA ODNIESIENIA 0 °C***Cold junction*

WIKA	głębokość zanurzenia = 150 mm	L-Z-0005
------	-------------------------------	----------

*Immersion depth***WARUNKI ŚRODOWISKOWE***Environmental conditions*

Temperatura otoczenia (21,2 ± 23,5) °C

Ambient temperature

Wilgotność względna (50,1 ± 58,7) %

Humidity

ŚWIADECTWO WZORCOWANIA NR: PA2504047/2

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201


Technologie pomiaru temperatury
Niezawodne - Dokładne - Certyfikowane

WYNIKI WZORCOWANIA

Calibration results

CZUJNIK TYPU Sensor type	NOMINALNA Nominal	TEMPERATURA Temperatura ODNIESIENIA Reference	OBLICZONA Calculated	ZMIERZONE NAPIĘCIE TERMIELEKTRYCZNE Thermovoltage	BŁĄD POMIARU Measurement error	NAJWIĘKSZY BŁĄD DOPUSZCZALNY Maximum permissible error	NIEPENNOŚĆ POMIARU Measurement uncertainty	UWAGI Comments
		[°C]		[mV]	[°C]	± [°C]	± [°C]	
K	10,0	10,0	10,0	0,398	0,0	1,5	0,7	
	58,0	57,8	57,4	2,326	-0,4	1,5	0,7	
	250,0	248,8	248,7	10,099	-0,1	1,5	1,2	
	500,0	499,5	499,5	20,624	0,0	2,0	1,2	
	950,0	949,9	948,8	39,268	-1,1	3,8	1,7	

UWAGI

Comments

- Temperatura nominalna: punkt nastawy wynikający z zamówienia.
Nominal temperature: set point specified in the order.
- Temperatura odniesienia: średnia wartość z pomiarów odczytanych z wzorca.
Reference temperature: average value of the measurement results read from the reference standard.
- Temperatura obliczona: wartość temperatury obliczona na podstawie normy ASTM E230M-23a z wartości zmierzonej napięcia termoelektrycznego obiektu.
Calculated temperature: temperature value calculated from the measured thermovoltage of the calibrated object on the basis of the ASTM E230M-23a.
- Błąd pomiaru = Temperatura obliczona - Temperatura odniesienia
Measurement error = Calculated temperature - Reference temperature
- Największy błąd dopuszczalny: dopuszczalny błąd pomiaru temperatury według normy klasy 1 wg PN EN 60584-1:2014 pkt 5.1.
Maximum permissible error: according to the Class 1 acc. to PN EN 60584-1: 2014 point 5.1.
- Stwierdzenie zgodności z daną specyfikacją podjęto na podstawie zasady prostej akceptacji ILAC G8:09/2019 w=0 TUR>3:1.
Ryzyko specyficzne błędnej akceptacji/ błędnego odrzucenia wynosi 50%.
- Compliance with the principles of the simple acceptance according to the ILAC G8:09/2019 w=0 TUR >3:1. The specific risk of a false acceptance/false rejection can be up to 50%.

ZGODNOŚĆ Z WYMAGANIAMI

Conformity with requirements

Na podstawie wszystkich wyników wzorcowania stwierdzono zgodność z wymaganiami PN-EN 60584-1:2014-04 pkt 5.1 dla klasy 1. Conformance with PN-EN 60584-1:2014-04 pt 5.1 class 1 was found on the basis of all calibration results.

Wykonał:
Made by

Podpis:
Signature

Pracownik Techniczny Laboratorium
Laboratory Technician
Paweł Krysztolik

Autoryzował:
Approved

Podpis:
Signature

Z-ca Kierownika Laboratorium
Deputy Head of Laboratory

mgr Joanna Zajac

KONIEC/ End

Data wzorcowania: 12.04.2025
Calibration date

Strona 3 / 3
Page

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 D Tel.: +48 (0)71 / 352 70 70
55-085 Dąbówka Faks: +48 (0)71 / 352 70 71
Polska biuro@guenther.com.pl - www.guenther.com.pl

NIP 893-19-06-889
REGON 020702521
KRS 0000352821

BNP PARIBAS BANK POLSKA S.A.
PLN 36 1750 1064 0000 0000 3055 6722
EUR 61 1750 1064 0000 0000 2055 6757 Kod SWIFT/BIC: PBPAPL33XXX




AP 201



Guenther Polska Sp. z o.o.
ul. Wrocławska 27C 55-095 Długołęka
Laboratorium wzorcowujące akredytowane
przez Polskie Centrum Akredytacji,
sygnatariusza porozumień EA MLA i ILAC MRA
dotyczących wzajemnego uznawania świadectw
Nr akredytacji AP 201
Calibration laboratory accredited by
Polish Centre for Accreditation,
a signatory to EA MLA and ILAC MRA
that include recognition of calibration certificates.
Accreditation No. AP 201



Technologia pomiaru temperatury
Niezawodna · Dokładna · Certyfikowana

ŚWIADECTWO WZORCOWANIA NR: PA2504112

CALIBRATION CERTIFICATE NO.

Data wydania: 30.04.2025

Release date

ZGLASZAJĄCY

Applicant

BALTIC FIRE LABORATORY Sp. z o.o.

ul. Banińska 8
80-209 Tuchom

MIEJSCE
WZORCOWANIA
Place of calibration

Guenther Polska Sp. z o.o.

ul. Wrocławska 27c
55-095 Długołęka

NR ZAMÓWIENIA ZEW.

External order No.

zam. z dnia 17.04.25

NR ZAMÓWIENIA WEW.

Internal order No.

ZS 233/04/2025
OBIEKT
WZORCOWANIA
Calibration object

Nazwa:

Name

Czujnik termoelektryczny

Thermocouple

Non-expendable

Ø : 1,5 mm

Typ:

Type

1x NiCr-NiAl/K

Kod artykułu:

Article code

20053105

Producent/Model:

Manufacturer/Model

Guenther

Szczeta:

Batch No.

126896
DATA WZORCOWANIA

Calibration date

29-30.04.2025

RODZAJ
WZORCOWANIA
Calibration type

Rekalibracja / Recalibration

METODA
WZORCOWANIA
Calibration method

Wzorcowanie zostało przeprowadzone zgodnie z procedurą QM/V9.01.01 Wzorcowanie termoelementu metodą porównawczą ver. 3.0 z dnia 12.10.2021 (opracowanej na podstawie ASTM E220-19).

The calibration has been performed in accordance with procedure QM/V9.01.01 Calibration of the thermocouple by the comparison method ver.3.0 of 12.10.2021 (based on the ASTM E220-19).

SPÓJNOŚĆ
POMIAROWA

Measurement traceability

Świadectwo potwierdza spójność wyników pomiarów z jednostkami miar Międzynarodowego Układu Jednostek Miar (SI).

The certificate provides traceability of measurement results to the International System of Units (SI).

NIEPEWNOŚĆ
POMIARU

Measurement uncertainty

Niepewność pomiaru została określona zgodnie z dokumentem EA-4/02 M:2022.

Podane wartości niepewności stanowią niepewności rozszerzone przy prawdopodobieństwie rozszerzenia ok. 95% i współczynnika rozszerzenia $k = 2$.

The measurement uncertainty has been determined in accordance with EA-4/02 M: 2022. The reported uncertainties of measurement are stated as the expanded uncertainties corresponded to a coverage probability of approximately 95% and a coverage factor $k = 2$.

UWAGI

Comments

Podane wyniki wzorcowania dotyczą wyłącznie wzorcowanego obiektu.

Calibration results apply only calibrated object.

W razie wątpliwości ważny jest oryginalny polski tekst.

If there are any doubts, the Polish version shall prevail.

Niniejsze świadectwo może być okazywane lub kopiowane tylko w całości.

This certificate may be presented or copied as a whole document only.

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 C Tel. +48 (0)71 / 362 70 70
55-095 Długołęka Fax: +48 (0)71 / 362 70 71
Polska biuro@guenther.com.pl - www.guenther.com.pl

NIP 886-18-06-589
REGON 020702531
KRS 0000308821

BNP PARIBAS BANK POLSKA S.A.
PLN 36 1750 1084 0000 0000 2055 6722
EUR 61 1750 1084 0000 0000 2055 6757 Kod SWIFT/BIC: PPARPL30XXX

Strona 1 / 5
Page



ŚWIADECTWO WZORCOWANIA NR: PA2504112

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

UŻYTE WZORCE I URZĄDZENIA POMIAROWE

Reference probes and measuring devices

WZORZEC

Reference probes

T ≤ 250 °C	Pt100 / 4	SW 45/35/INTIBS-LWT/2025	L-R-0026	07-04-2026
T > 250 °C	Pt100 / 4	4166/2024	L-R-0021	28-10-2025
T > 500 °C	PtRh10-Pt/S	PA2411069/1	L-T-0040	19-11-2025

URZĄDZENIE POMIAROWE

Measuring instrument

T ≤ 58 °C	Keysight	2672/2024	L-M-0007	14-07-2025
T > 58 °C	Keysight	4990/2024	L-M-0009	22-12-2025
T > 500 °C	Keysight	76/2025	L-M-0008	16-01-2026

ELEMENT GRZEJNY

Temperature source

T ≤ 58 °C	Fluke	głębokość zanurzenia = 150 mm	L-H-0002
T > 58 °C	Fluke	głębokość zanurzenia = 150 mm	L-H-0003
T > 500 °C	Additel	głębokość zanurzenia = 150 mm	L-H-0011

Immersion depth

SPOINA ODNIESIENIA 0 °C

Cold junction

T ≤ 500 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0002
T > 500 °C	WIKA	głębokość zanurzenia = 150 mm	L-Z-0005

Immersion depth

WARUNKI ŚRODOWISKOWE

Environmental conditions

Temperatura otoczenia	(22,4 ± 24,7) °C
Ambient temperature	
Wilgotność względna	(28,2 ± 30,5) %
Humidity	

ŚWIADECTWO WZORCOWANIA NR: PA2504112

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201


WYNIKI WZORCOWANIA
Results of calibration

CZUJNIK Typu Sensor type	NOMINALNA Nominal	TEMPERATURA Temperatura ODWIECZENIA Reference	OBLICZONA Calculated	ZMIERZONE NAPIĘCIE TERMOWELEKTRYCZNE Thermovoltage	BŁĄD POMIARU Measurement error	NAJWIĘKSZY BŁĄD DOPUSZCZALNY Maximum permissible error ±/°C	WIECZYSTOŚĆ POMIARU Measurement uncertainty ±/°C	UWAGI Comments
		[°C]		[mV]	[°C]			
K	10,0	10,1	9,9	0,396	-0,2	1,5	0,7	1831220/1 PA2504112/1
	58,0	57,9	57,6	2,338	-0,3	1,5	0,7	
	250,0	250,0	249,8	10,144	-0,2	1,5	0,7	
	500,0	500,4	500,2	20,655	-0,2	2,0	0,7	
	950,0	950,2	950,2	39,323	0,0	3,8	1,8	
K	10,0	10,1	9,9	0,395	-0,2	1,5	0,7	1831220/14 PA2504112/2
	58,0	57,9	57,7	2,340	-0,2	1,5	0,7	
	250,0	250,0	249,8	10,146	-0,2	1,5	0,7	
	500,0	500,4	499,9	20,642	-0,5	2,0	0,7	
	950,0	950,1	950,2	39,323	0,1	3,8	1,8	
K	10,0	10,1	9,9	0,396	-0,2	1,5	0,7	1831220/17 PA2504112/3
	58,0	57,9	57,6	2,336	-0,3	1,5	0,7	
	250,0	250,0	249,4	10,129	-0,6	1,5	0,7	
	500,0	500,4	499,5	20,624	-0,9	2,0	0,7	
	950,0	950,2	949,3	39,286	-0,9	3,8	1,8	
K	10,0	10,0	9,8	0,391	-0,2	1,5	0,7	1831220/18 PA2504112/4
	58,0	57,9	57,6	2,337	-0,3	1,5	0,7	
	250,0	250,1	249,5	10,131	-0,6	1,5	0,7	
	500,0	500,3	499,6	20,626	-0,7	2,0	0,7	
	950,0	950,2	949,8	39,304	-0,4	3,8	1,8	
K	10,0	10,0	9,9	0,396	-0,1	1,5	0,7	1831220/25 PA2504112/5
	58,0	57,9	57,5	2,331	-0,4	1,5	0,7	
	250,0	250,1	248,4	10,089	-1,7	1,5	0,7	
	500,0	500,5	497,5	20,536	-3,0	2,0	0,7	
	950,0	950,0	947,4	39,210	-2,6	3,8	1,8	
K	10,0	10,0	9,8	0,392	-0,2	1,5	0,7	1831220/28 PA2504112/6
	58,0	57,9	57,6	2,337	-0,3	1,5	0,7	
	250,0	250,1	249,6	10,137	-0,5	1,5	0,7	
	500,0	500,3	499,4	20,618	-0,9	2,0	0,7	
	950,0	950,0	950,1	39,320	0,1	3,8	1,8	
K	10,0	10,0	9,8	0,390	-0,2	1,5	0,7	1831220/32 PA2504112/7
	58,0	57,9	57,8	2,343	-0,1	1,5	0,7	
	250,0	250,0	250,3	10,166	0,3	1,5	0,7	
	500,0	500,6	500,9	20,686	0,3	2,0	0,7	
	950,0	950,0	950,5	39,333	0,5	3,8	1,8	

Data wzorcowania: 29-30.04.2025

Calibration date

Strona 3 / 5

Page

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 C Tel. +48 (0)71 / 352 70 70
55-065 Drugolęka Faks +48 (0)71 / 352 70 71
Polska biuro@guenther.com.pl - www.guenther.com.pl

NIP 895-19-06-589
REGON 020702531
KRS 000030821

BNP PARIBAS BANK POLSKA S.A.
PLN 36 1750 1064 0000 0000 2055 5722
ELR 61 1750 1064 0000 0000 2055 6757 Kod SWIFT/BIC: PABSPKXXXX


ŚWIADECTWO WZORCOWANIA NR: PA2504112

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

CZUJNIK TYPU Sensor type	TEMPERATURA NAPIĘCIE CONVERSION Reference			OBLICZONA Calculated	ZMIERZONE BŁĄD TERMIELEKTRYCZNE Thermovoltage	BŁĄD POMIARU Measurement error	NAJWIĘKSZY BŁĄD DOPUSZCZALNY Maximum permissible error	NIEPEWNOŚĆ POMIARU Measurement uncertainty	UWAGI Comments
	NOMINALNA Nominal								
	[°C]				[mV]	[°C]	±[°C]	±[°C]	
K	10,0	10,1	9,9		0,394	-0,2	1,5	0,7	1831220/35 PA2504112/8
	58,0	57,9	57,7		2,339	-0,2	1,5	0,7	
	250,0	250,0	249,8		10,145	-0,2	1,5	0,7	
	500,0	500,4	500,0		20,646	-0,4	2,0	0,7	
	950,0	950,2	950,0		39,313	-0,2	3,8	1,8	
K	10,0	10,0	9,8		0,392	-0,2	1,5	0,7	1831220/36 PA2504112/9
	58,0	57,9	57,7		2,338	-0,2	1,5	0,7	
	250,0	250,1	249,5		10,135	-0,6	1,5	0,7	
	500,0	500,3	499,3		20,617	-1,0	2,0	0,7	
	950,0	950,2	949,8		39,306	-0,4	3,8	1,8	
K	10,0	10,0	9,8		0,390	-0,2	1,5	0,7	1831220/39 PA2504112/10
	58,0	57,9	57,7		2,342	-0,2	1,5	0,7	
	250,0	250,1	250,2		10,162	0,1	1,5	0,7	
	500,0	500,3	500,3		20,658	0,0	2,0	0,7	
	950,0	950,1	950,8		39,344	0,7	3,8	1,8	

UWAGI

Comments

- Temperatura nominalna: punkt nastawy wynikający z zamówienia.
Nominal temperature: set point specified in the order.
- Temperatura odniesienia: średnia wartość z pomiarów odczytanych z wzorca.
Reference temperature: average value of the measurement results read from the reference standard.
- Temperatura obliczona: wartość temperatury obliczona na podstawie normy ASTM E230M-23a z wartości zmierzonej napięcia termoelektrycznego obiektu.
Calculated temperature: temperature value calculated from the measured thermovoltage of the calibrated object on the basis of the ASTM E230M-23a.
- Błąd pomiaru = Temperatura obliczona - Temperatura odniesienia
Measurement error = Calculated temperature - Reference temperature
- Największy błąd dopuszczalny: dopuszczalny błąd pomiaru temperatury według normy klasy 1 wg PN EN 60584-1:2014 pkt 5.1.
Maximum permissible error: according to the Class 1 acc. to PN EN 60584-1: 2014 point 5.1.
- Stwierdzenie zgodności z daną specyfikacją podjęto na podstawie zasady prostej akceptacji ILAC G8:09/2019 w=0 TUR>3:1.
- Ryzyko specyficzne błędnej akceptacji/ błędnego odrzucenia wynosi 50%.
- Compliance with the principles of the simple acceptance according to the ILAC G8:09/2019 w=0 TUR >3:1. The specific risk of a false acceptance/false rejection can be up to 50%.

ŚWIADECTWO WZORCOWANIA NR: PA2504112

CALIBRATION CERTIFICATE NO.

Wydane przez LABORATORIUM AKREDYTOWANE Nr AP 201

Released by Accredited Laboratory No. AP 201

**ZGODNOŚĆ Z WYMAGANIAMI***Conformity with requirements*

Na podstawie wszystkich wyników wzorcowania dla czujników o numerach 1831220/1, 1831220/14, 1831220/17, 1831220/18, 1831220/28, 1831220/32, 1831220/35, 1831220/36, 1831220/39 stwierdzono zgodność z wymaganiami PN-EN 60584-1:2014-04 pkt 5.1 dla klasy 1. Conformance with PN-EN 60584-1:2014-04 pt 5.1 class 1 was found on the basis of all calibration results for thermocouples No. 1831220/1, 1831220/14, 1831220/17, 1831220/18, 1831220/28, 1831220/32, 1831220/35, 1831220/36, 1831220/39.

Na podstawie wyników wzorcowania dla czujnika o numerze 1831220/25 w zakresie $<10^{\circ}\text{C}$; 58°C oraz w punkcie pomiarowym 950°C stwierdzono zgodność z wymaganiami PN-EN 60584-1:2014-04 pkt 5.1. Natomiast w zakresie $<250^{\circ}\text{C}$; 500°C stwierdzono brak zgodności z wymaganiami PN-EN 60584-1:2014-04 pkt 5.1. dla tego czujnika. Conformance with PN-EN 60584-1:2014-04 was found on the basis of calibration results for thermocouple No. 1831220/25 in the range $<10^{\circ}\text{C}$; 58°C and at calibration point 950°C . Nonconformance with PN-EN 60584-1:2014-04 pt 5.1 was found for this thermocouple on the basis of calibration results range $<250^{\circ}\text{C}$; 500°C .

Wykonał: A.Grykałowska

Made by

Inżynier Laboratorium
Laboratory Engineer

dr inż. Agnieszka Grykałowska

Autoryzował:

Approved

Z-ca Kierownika Laboratorium

Deputy Head of Laboratory

Podpis:

Signature

mgr Joanna Zajac

KONIEC/ End

Data wzorcowania: 29-30.04.2025

Calibration date

Strona 5 / 5

Page

GUENTHER Polska Sp. z o.o.

ul. Wrocławska 27 C
35-085 Olędzka
Polska

Tel. +48 (0)71 / 952 70 70

Faks +48 (0)71 / 952 70 71

biuro@guenther.com.pl - www.guenther.com.pl

NIP 885-19-06-588

REGON 020702531

KRS 0000300821

BNP PARIBAS BANK POLSKA S.A.

PLN 36 1750 1064 0000 0000 2035 6722

EUR 61 1750 1064 0000 0000 2035 6757 Kod SWIFT/BIC: PRABPL30XXX



DNV



Appendix F – laboratory certificates

POLSKIE CENTRUM AKREDYTACJI
POLISH CENTRE FOR ACCREDITATION



Sygnatariusz EA MLA
EA MLA Signatory

CERTYFIKAT AKREDYTACJI
LABORATORIUM BADAWCZEGO
ACCREDITATION CERTIFICATE OF TESTING LABORATORY
Nr AB 1792

Potwierdza się, że: / This is to confirm that:

BALTIC FIRE LABORATORY Sp. z o.o.
BFL FIRE TESTING LABORATORY
ul. Banińska 8
80-209 Tuchom


spełnia wymagania normy PN-EN ISO/IEC 17025:2018-02
meets requirements of the PN-EN ISO/IEC 17025:2018-02 standard

Akredytowana działalność jest określona w Zakresie Akredytacji Nr AB 1792
Accredited activity is defined in the Scope of Accreditation No AB 1792

Akredytacja pozostaje w mocy pod warunkiem przestrzegania
wymagań jednostki akredytującej określonych w kontrakcie Nr AB 1792
This accreditation remains in force provided the Laboratory observes
the requirements of Accreditation Body defined in the Contract No AB 1792



DYREKTOR
POLSKIEGO CENTRUM AKREDYTACJI



LUCYNA OLBORSKA

Warszawa, 9 lutego 2022 roku

PCA

Zakres akredytacji Nr AB 1792
Scope of accreditation No AB 1792

**ZAKRES AKREDYTACJI
LABORATORIUM BADAWCZEGO**
SCOPE OF ACCREDITATION FOR TESTING LABORATORY
Nr/No AB 1792

wydany przez / issued by
POLSKIE CENTRUM AKREDYTACJI
01-382 Warszawa, ul. Szczotkarska 42

Wydanie/Issue 7 z/of 05.09.2025

 AB 1792	Nazwa i adres / Name and address BALTIC FIRE LABORATORY Sp. z o.o. BFL FIRE TESTING LABORATORY ul. Banińska 8 80-209 Tuchom
Kod identyfikacyjny / Identification code *) - H/17	Dziedzina i przedmiot badań / Field of testing and item: - Badania ogniowe wyrobów innych / Fire tests of other products

Wersja strony/Page version: A

⁷⁾ Kod identyfikacyjny zgodnie z załącznikiem do dokumentu DAB-07 dostępnym na stronie internetowej www.pca.gov.pl /
The identification code according to the Annex to document DAB-07, available at PCA website www.pca.gov.pl



**KIEROWNIK DZIAŁU AKREDYTACJI
BADAŃ MECHANICZNYCH I FIZYCZNYCH**


MARIA SZAFRAN

Niniejszy dokument jest załącznikiem do Certyfikatu Akredytacji Nr AB 1792 z dnia 09.02.2022 r.
Cykl akredytacji od 11.07.2025 r. do 18.07.2029 r.
Status akredytacji oraz aktualność zakresu akredytacji można potwierdzić na stronie internetowej PCA www.pca.gov.pl

This document is an annex to accreditation certificate No AB 1792 of 09.02.2022
Accreditation cycle from 11.07.2025 to 18.07.2029

The status of accreditation and validity of the scope of accreditation can be confirmed at PCA website www.pca.gov.pl

Wydanie nr / Issue No 7, 05.09.2025 1/7

BFL Fire Testing Laboratory ul. Banińska 8, 80-209 Tuchom		
Przedmiot badań/wyrób	Rodzaj działalności/badane cechy/metoda	Dokumenty odniesienia
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia i tłumienia ognia Temperatura Stężenie (O ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas Obszar uszkodzenia, %	IMO Res. A.800(19) ze zmianami w IMO MSC.265(84) Appendix 2 PN-EN 14972-1:2021, Annex A PN-EN 14972-3:2022-01 PN-EN 14972-4:2024 PN-EN 14972-10:2022-09 UL 2167 (2021) pkt 44, 45, 46, 48, 49, 50 VdS 3883-1:2020-06 VdS 3883-2:2020-06 VdS 3883-5:2020-06 FM Approval Standard Class 5560:2021 Appendix P
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia ognia Temperatura Stężenie (O ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	IMO MSC/Circ.1165 Appendix B ze zmianami w IMO MSC.1/Circ.1269 oraz IMO MSC.1/Circ.1386 UL 2167 (2021) pkt 43 FM Approval Standard Class 5560:2021 Appendix E FM Approval Standard Class 5560:2021 Appendix F PN-EN 14972-8:2020-06 PN-EN 14972-14:2022-02 PN-EN 14972-15:2022-02 PN-EN 14972-9:2020-06
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia i tłumienia ognia Temperatura Stężenie (O ₂) Stężenie (CO) Stężenie (CO ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas Obszar uszkodzenia, %	PN-EN 14972-7:2024

Wersja strony: A

Przedmiot badań/wyrób	Rodzaj działalności/badane cechy/metoda	Dokumenty odniesienia
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia i tłumienia ognia Rozprzestrzenianie się płomienia Temperatura Stężenie (O ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	IMO MSC.1/Circ.1268
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia ognia Rozprzestrzenianie się płomienia Temperatura Stężenie (O ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas Zakres: t > 1 s	IMO MSC.1/Circ.1387 Appendix
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność tłumienia ognia Rozprzestrzenianie się płomienia Temperatura Stężenie tlenu Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas Obszar uszkodzenia, %	IMO MSC.1/Circ.1430/Rev.03 Appendix VdS 3883-4:2020-06 PN-EN 14972-5:2024
Wodne systemy przeciwpożarowe Materiały instalacyjne	Rozprzestrzenianie się płomienia Temperatura Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	A.753(18) GUIDELINES FOR THE APPLICATION OF PLASTIC PIPES ON SHIPS
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia ognia Rozprzestrzenianie się płomienia Temperatura Stężenie tlenu Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	FM Approval Standard Class 5560:2021 Appendix M FM Approval Standard Class 5560:2021 Appendix N
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia i tłumienia ognia Temperatura Stężenie tlenu Objętościowe natężenie przepływu (metoda wagowa, metoda bezpośrednia) Ciśnienie Czas Obszar uszkodzenia, %	VdS 3883-3:2020-06 PN-EN 14972-6:2023-10

Wersja strony: A

Przedmiot badań/wyrób	Rodzaj działalności/badane cechy/metoda	Dokumenty odniesienia
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia i tłumienia ognia Temperatura Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	VdS 3883-6:2020-06 BS EN 14972-17:2025 UL 2167 (2021) pkt 47 UL 2167A (2024) pkt 23
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia i tłumienia ognia Temperatura Stężenie (O ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas Obszar uszkodzenia, %	VdS 3883-7:2020-06
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia i tłumienia ognia Temperatura Stężenie (O ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas Obszar uszkodzenia, %	VdS 3883-8:2020-06 PN EN 14972-11:2023-10
Aerosolowe systemy przeciwpożarowe Generatory aerosolu	Skuteczność gaszenia ognia Temperatura Stężenie (O ₂) Stężenie (CO) Stężenie (CO ₂) Ciśnienie (Pa) Pomiar masy Czas	IMO MSC.1/Circ.1270 ISO 15779 Annex D
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia i tłumienia ognia Temperatura Stężenie (O ₂) Stężenie (CO) Stężenie (CO ₂) Stężenie (HCN) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	ARGE Guideline - Part 2:2018 Rev. 4.1 pkt. 3.2, 3.3 UNI 11565:2021 Appendix B
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność tłumienia ognia Temperatura Stężenie tlenu Stężenie (O ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas Stopień uszkodzenia, %	Procedura badawcza BFL TP-02 ver. 4 z 13.08.2021 PN-EN 14972-1:2021, Annex A

Wersja strony: A

Przedmiot badań/wyrób	Rodzaj działalności/badane cechy/metoda	Dokumenty odniesienia
Gazowe systemy przeciwpożarowe	Skuteczność gaszenia ognia Temperatura Stężenie (O ₂) Stężenie (CO) Stężenie (CO ₂) Ciśnienie (Pa) Pomiar masy Czas	NCC71 - 1.1.2021 Chapter 2
Gazowe systemy przeciwpożarowe	Skuteczność gaszenia ognia Temperatura Stężenie (O ₂) Ciśnienie (Pa) Pomiar masy Czas	ISO 14520-1:2023 Annex C UL2166 (2021) pkt 35, 36 ,37 FM Approval Standard Class 5600:2023 Chapter 4.1
Pianowe systemy przeciwpożarowe Koncentraty pianowe	Skuteczność gaszenia ognia Temperatura Czas	MSC.1/Circ.1312 pkt 3.9 UL 162 (2021) pkt 11, 12 FM Approval Standard Class 5130:2021 chapter 4.2, 4.3, 4.7
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia ognia Temperatura Stężenie (O ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	EN 14972-12:2023
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Temperatura Ciśnienie	IMO MSC/Circ.1165 pkt 4.10
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia ognia Temperatura Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	ISO 15371:2015 PN-EN 17446:2021
Wodne systemy przeciwpożarowe Dysze tryskaczowe	Skuteczność gaszenia ognia Temperatura Stężenie (O ₂) Stężenie (CO) Stężenie (CO ₂) Objętościowe natężenie przepływu (metoda bezpośrednia, metoda wagowa) Ciśnienie Czas	FM Approval Standard Class 5560:2021 Appendix A FM Approval Standard Class 5560:2021 Appendix B FM Approval Standard Class 5560:2021 Appendix C FM Approval Standard Class 5560:2021 Appendix D FM Approval Standard Class 5560:2021 Appendix I JH-C077-E-RP-001
Gaśnice przenośne	Minimalna skuteczność gaśnicza – ocena pożaru testowego grupy A, Minimalna skuteczność gaśnicza – ocena pożaru testowego grupy B	PN-EN 3-7:2008 (Appendix I)

Wersja strony: A

PCA

Zakres akredytacji Nr AB 1792

Przedmiot badań/wyrób	Rodzaj działalności/badane cechy/metoda	Dokumenty odniesienia
Gaśnice przewoźne	Minimalna skuteczność gaśnicza – ocena pożaru testowego grupy A Minimalna skuteczność gaśnicza – ocena pożaru testowego grupy B	PN-EN 1866-1:2010 (chapter 8)

Wersja strony: A

Wykaz zmian Zakresu Akredytacji Nr AB 1792

Status zmian: wersja pierwotna - A



Zatwierdzam status zmian
KIEROWNIK DZIAŁU AKREDYTACJI
BADAŃ MECHANICZNYCH I FIZYCZNYCH


MARIA SZAFRAN
dnia: 05.09.2025 r.

TEST REPORT END